

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2458

Analytical Method: 8081B

Client: CDM Smith

Datafile : PD089330.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB168718BS (Column 1)	alpha-BHC	0.5	0.56	ug/L	111				85	130	
	beta-BHC	0.5	0.53	ug/L	106				83	126	
	delta-BHC	0.5	0.59	ug/L	117				69	141	
	gamma-BHC (Lindane)	0.5	0.55	ug/L	110				82	129	
	Heptachlor	0.5	0.54	ug/L	107				79	127	
	Aldrin	0.5	0.56	ug/L	113				79	126	
	Heptachlor epoxide	0.5	0.55	ug/L	110				81	124	
	Endosulfan I	0.5	0.56	ug/L	112				85	122	
	Dieldrin	0.5	0.57	ug/L	113				83	125	
	4,4'-DDE	0.5	0.56	ug/L	112				80	127	
	Endrin	0.5	0.49	ug/L	98				81	128	
	Endosulfan II	0.5	0.57	ug/L	113				82	123	
	4,4'-DDD	0.5	0.59	ug/L	118				77	131	
	Endosulfan sulfate	0.5	0.55	ug/L	110				76	129	
	4,4'-DDT	0.5	0.48	ug/L	96				80	133	
	Methoxychlor	0.5	0.42	ug/L	83				78	108	
	Endrin ketone	0.5	0.56	ug/L	111				80	131	
	Endrin aldehyde	0.5	0.56	ug/L	112				82	127	
	alpha-Chlordane	0.5	0.56	ug/L	111				82	125	
	gamma-Chlordane	0.5	0.57	ug/L	114				82	125	
PB168718BS (Column 2)	alpha-BHC	0.5	0.56	ug/L	112				85	130	
	beta-BHC	0.5	0.54	ug/L	109				83	126	
	delta-BHC	0.5	0.55	ug/L	110				69	141	
	gamma-BHC (Lindane)	0.5	0.56	ug/L	111				82	129	
	Heptachlor	0.5	0.52	ug/L	105				79	127	
	Aldrin	0.5	0.56	ug/L	111				79	126	
	Heptachlor epoxide	0.5	0.55	ug/L	111				81	124	
	Endosulfan I	0.5	0.55	ug/L	111				85	122	
	Dieldrin	0.5	0.55	ug/L	110				83	125	
	4,4'-DDE	0.5	0.55	ug/L	110				80	127	
	Endrin	0.5	0.48	ug/L	95				81	128	
	Endosulfan II	0.5	0.55	ug/L	110				82	123	
	4,4'-DDD	0.5	0.56	ug/L	112				77	131	
	Endosulfan sulfate	0.5	0.55	ug/L	109				76	129	
	4,4'-DDT	0.5	0.47	ug/L	95				80	133	
	Methoxychlor	0.5	0.44	ug/L	88				78	108	
	Endrin ketone	0.5	0.55	ug/L	110				80	131	
	Endrin aldehyde	0.5	0.55	ug/L	110				82	127	
	alpha-Chlordane	0.5	0.55	ug/L	109				82	125	
	gamma-Chlordane	0.5	0.55	ug/L	109				82	125	