

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>CAMP02</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q2458</u>
Instrument ID:	<u>BNA_F</u>	Calibration Date/Time:	<u>07/01/2025 10:47</u>
Lab File ID:	<u>BF142941.D</u>	Init. Calib. Date(s):	<u>06/19/2025 06/19/2025</u>
EPA Sample No.:	<u>SSTDCCC040</u>	Init. Calib. Time(s):	<u>17:07 20:40</u>
GC Column:	<u>DB-UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
2-Fluorophenol	1.201	1.196		-0.4	
Benzaldehyde	0.841	0.981		16.6	
Phenol-d6	1.417	1.371		-3.2	
Phenol	1.575	1.539		-2.3	20.0
bis(2-Chloroethyl)ether	1.126	1.074		-4.6	
2-Chlorophenol	1.296	1.271		-1.9	
2-Methylphenol	0.982	0.943		-4.0	
2,2-oxybis(1-Chloropropane)	1.809	1.620		-10.4	
Acetophenone	0.441	0.440		-0.2	
3+4-Methylphenols	1.224	1.200		-2.0	
n-Nitroso-di-n-propylamine	0.824	0.752	0.050	-8.7	
Nitrobenzene-d5	0.365	0.388		6.3	
Hexachloroethane	0.511	0.499		-2.3	
Nitrobenzene	0.330	0.317		-3.9	
Isophorone	0.585	0.543		-7.2	
2-Nitrophenol	0.180	0.182		1.1	20.0
2,4-Dimethylphenol	0.311	0.298		-4.2	
bis(2-Chloroethoxy)methane	0.372	0.351		-5.6	
2,4-Dichlorophenol	0.282	0.278		-1.4	20.0
Naphthalene	0.979	0.962		-1.7	
4-Chloroaniline	0.398	0.373		-6.3	
Hexachlorobutadiene	0.190	0.192		1.1	20.0
Caprolactam	0.075	0.066		-12.0	
4-Chloro-3-methylphenol	0.281	0.259		-7.8	20.0
2-Methylnaphthalene	0.607	0.581		-4.3	
Hexachlorocyclopentadiene	0.335	0.285	0.050	-14.9	
2,4,6-Trichlorophenol	0.385	0.386		0.3	20.0
2-Fluorobiphenyl	1.522	1.666		9.5	
2,4,5-Trichlorophenol	0.405	0.387		-4.4	
1,1-Biphenyl	1.580	1.580		0.0	
2-Chloronaphthalene	1.186	1.191		0.4	
2-Nitroaniline	0.332	0.310		-6.6	
Dimethylphthalate	1.295	1.215		-6.2	
Acenaphthylene	1.952	1.921		-1.6	
2,6-Dinitrotoluene	0.284	0.270		-4.9	
3-Nitroaniline	0.313	0.286		-8.6	
Acenaphthene	1.191	1.163		-2.4	20.0
2,4-Dinitrophenol	0.142	0.118	0.050	-16.9	
4-Nitrophenol	0.214	0.170	0.050	-20.6	

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>CAMP02</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q2458</u>
Instrument ID:	<u>BNA_F</u>	Calibration Date/Time:	<u>07/01/2025 10:47</u>
Lab File ID:	<u>BF142941.D</u>	Init. Calib. Date(s):	<u>06/19/2025 06/19/2025</u>
EPA Sample No.:	<u>SSTDCCC040</u>	Init. Calib. Time(s):	<u>17:07 20:40</u>
GC Column:	<u>DB-UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
Dibenzofuran	1.708	1.646		-3.6	
2,4-Dinitrotoluene	0.378	0.345		-8.7	
Diethylphthalate	1.269	1.171		-7.7	
4-Chlorophenyl-phenylether	0.636	0.605		-4.9	
Fluorene	1.334	1.252		-6.1	
4-Nitroaniline	0.286	0.264		-7.7	
4,6-Dinitro-2-methylphenol	0.121	0.115		-5.0	
n-Nitrosodiphenylamine	0.693	0.694		0.1	20.0
2,4,6-Tribromophenol	0.206	0.185		-10.2	
4-Bromophenyl-phenylether	0.228	0.233		2.2	
Hexachlorobenzene	0.253	0.254		0.4	
Atrazine	0.179	0.187		4.5	
Pentachlorophenol	0.126	0.106		-15.9	20.0
Phenanthrene	1.083	1.062		-1.9	
Anthracene	1.111	1.107		-0.4	
Carbazole	0.959	0.938		-2.2	
Di-n-butylphthalate	0.999	1.056		5.7	
Fluoranthene	1.028	1.022		-0.6	20.0
Pyrene	1.875	1.566		-16.5	
Terphenyl-d14	1.447	1.189		-17.8	
Butylbenzylphthalate	0.544	0.580		6.6	
3,3-Dichlorobenzidine	0.438	0.458		4.6	
Benzo (a) anthracene	1.349	1.288		-4.5	
Chrysene	1.204	1.192		-1.0	
Bis(2-ethylhexyl)phthalate	0.782	0.932		19.2	
Di-n-octyl phthalate	1.475	1.697		15.1	20.0
Benzo (b) fluoranthene	1.157	1.116		-3.5	
Benzo (k) fluoranthene	1.083	1.089		0.6	
Benzo (a) pyrene	1.118	1.105		-1.2	20.0
Indeno (1,2,3-cd) pyrene	1.523	1.413		-7.2	
Dibenzo (a,h) anthracene	1.238	1.151		-7.0	
Benzo (g,h,i) perylene	1.234	1.133		-8.2	
1,2,4,5-Tetrachlorobenzene	0.590	0.606		2.7	
1,4-Dioxane	0.480	0.455		-5.2	20.0
2,3,4,6-Tetrachlorophenol	0.326	0.292		-10.4	

All other compounds must meet a minimum RRF of 0.010.