

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>CAMP02</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q2458</u>
Instrument ID:	<u>BNA_F</u>	Calibration Date/Time:	<u>07/03/2025 09:55</u>
Lab File ID:	<u>BF142987.D</u>	Init. Calib. Date(s):	<u>06/19/2025 06/19/2025</u>
EPA Sample No.:	<u>SSTDCCC040</u>	Init. Calib. Time(s):	<u>17:07 20:40</u>
GC Column:	<u>DB-UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
2-Fluorophenol	1.201	1.155		-3.8	
Benzaldehyde	0.841	0.872		3.7	
Phenol-d6	1.417	1.350		-4.7	
Phenol	1.575	1.488		-5.5	20.0
bis(2-Chloroethyl)ether	1.126	1.068		-5.2	
2-Chlorophenol	1.296	1.258		-2.9	
2-Methylphenol	0.982	0.942		-4.1	
2,2-oxybis(1-Chloropropane)	1.809	1.607		-11.2	
Acetophenone	0.441	0.435		-1.4	
3+4-Methylphenols	1.224	1.205		-1.6	
n-Nitroso-di-n-propylamine	0.824	0.783	0.050	-5.0	
Nitrobenzene-d5	0.365	0.383		4.9	
Hexachloroethane	0.511	0.494		-3.3	
Nitrobenzene	0.330	0.319		-3.3	
Isophorone	0.585	0.556		-5.0	
2-Nitrophenol	0.180	0.181		0.6	20.0
2,4-Dimethylphenol	0.311	0.288		-7.4	
bis(2-Chloroethoxy)methane	0.372	0.360		-3.2	
2,4-Dichlorophenol	0.282	0.274		-2.8	20.0
Naphthalene	0.979	0.960		-1.9	
4-Chloroaniline	0.398	0.386		-3.0	
Hexachlorobutadiene	0.190	0.191		0.5	20.0
Caprolactam	0.075	0.071		-5.3	
4-Chloro-3-methylphenol	0.281	0.270		-3.9	20.0
2-Methylnaphthalene	0.607	0.585		-3.6	
Hexachlorocyclopentadiene	0.335	0.268	0.050	-20.0	
2,4,6-Trichlorophenol	0.385	0.379		-1.6	20.0
2-Fluorobiphenyl	1.522	1.626		6.8	
2,4,5-Trichlorophenol	0.405	0.404		-0.2	
1,1-Biphenyl	1.580	1.566		-0.9	
2-Chloronaphthalene	1.186	1.170		-1.3	
2-Nitroaniline	0.332	0.323		-2.7	
Dimethylphthalate	1.295	1.278		-1.3	
Acenaphthylene	1.952	1.899		-2.7	
2,6-Dinitrotoluene	0.284	0.278		-2.1	
3-Nitroaniline	0.313	0.303		-3.2	
Acenaphthene	1.191	1.172		-1.6	20.0
2,4-Dinitrophenol	0.142	0.131	0.050	-7.7	
4-Nitrophenol	0.214	0.226	0.050	5.6	

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>CAMP02</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q2458</u>
Instrument ID:	<u>BNA_F</u>	Calibration Date/Time:	<u>07/03/2025 09:55</u>
Lab File ID:	<u>BF142987.D</u>	Init. Calib. Date(s):	<u>06/19/2025 06/19/2025</u>
EPA Sample No.:	<u>SSTDCCC040</u>	Init. Calib. Time(s):	<u>17:07 20:40</u>
GC Column:	<u>DB-UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
Dibenzofuran	1.708	1.633		-4.4	
2,4-Dinitrotoluene	0.378	0.365		-3.4	
Diethylphthalate	1.269	1.256		-1.0	
4-Chlorophenyl-phenylether	0.636	0.627		-1.4	
Fluorene	1.334	1.318		-1.2	
4-Nitroaniline	0.286	0.271		-5.2	
4,6-Dinitro-2-methylphenol	0.121	0.117		-3.3	
n-Nitrosodiphenylamine	0.693	0.695		0.3	20.0
2,4,6-Tribromophenol	0.206	0.200		-2.9	
4-Bromophenyl-phenylether	0.228	0.231		1.3	
Hexachlorobenzene	0.253	0.247		-2.4	
Atrazine	0.179	0.189		5.6	
Pentachlorophenol	0.126	0.112		-11.1	20.0
Phenanthrene	1.083	1.056		-2.5	
Anthracene	1.111	1.095		-1.4	
Carbazole	0.959	0.918		-4.3	
Di-n-butylphthalate	0.999	1.042		4.3	
Fluoranthene	1.028	0.963		-6.3	20.0
Pyrene	1.875	1.861		-0.7	
Terphenyl-d14	1.447	1.388		-4.1	
Butylbenzylphthalate	0.544	0.616		13.2	
3,3-Dichlorobenzidine	0.438	0.467		6.6	
Benzo (a) anthracene	1.349	1.360		0.8	
Chrysene	1.204	1.153		-4.2	
Bis(2-ethylhexyl)phthalate	0.782	0.921		17.8	
Di-n-octyl phthalate	1.475	1.699		15.2	20.0
Benzo (b) fluoranthene	1.157	1.162		0.4	
Benzo (k) fluoranthene	1.083	1.015		-6.3	
Benzo (a) pyrene	1.118	1.091		-2.4	20.0
Indeno (1,2,3-cd) pyrene	1.523	1.412		-7.3	
Dibenzo (a,h) anthracene	1.238	1.144		-7.6	
Benzo (g,h,i) perylene	1.234	1.104		-10.5	
1,2,4,5-Tetrachlorobenzene	0.590	0.594		0.7	
1,4-Dioxane	0.480	0.430		-10.4	20.0
2,3,4,6-Tetrachlorophenol	0.326	0.310		-4.9	

All other compounds must meet a minimum RRF of 0.010.