

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BF143024.D
Instrument ID: BNA_F

Contract: CAMP02
SDG NO.: Q2458
DFTPP Injection Date: 07/08/2025
DFTPP Injection Time: 10:24

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	34.6
68	Less than 2.0% of mass 69	0.6 (2) 1
69	Mass 69 relative abundance	32.1
70	Less than 2.0% of mass 69	0.2 (0.5) 1
127	10.0 - 80.0% of mass 198	44.5
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6
275	10.0 - 60.0% of mass 198	25.6
365	Greater than 1% of mass 198	3.3
441	Present, but less than mass 443	15.7
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	18.8 (18.8) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BF143025.D	07/08/2025	10:54
PB168716BL	PB168716BL	BF143026.D	07/08/2025	11:24
PB168716BS	PB168716BS	BF143027.D	07/08/2025	11:54
PB168716BSD	PB168716BSD	BF143028.D	07/08/2025	12:25