

Instrument ID: MSVOA\_X

**Daily Analysis Runlog For Sequence/QC Batch ID # VX070225**

|              |               |                   |                                   |
|--------------|---------------|-------------------|-----------------------------------|
| Review By    | John Carlone  | Review On         | 7/3/2025 8:44:30 AM               |
| Supervise By | Maresh Dadoda | Supervise On      | 7/3/2025 3:09:59 PM               |
| SubDirectory | VX070225      | HP Acquire Method | HP Processing Method 82X070225W.M |

| STD. NAME                                                                                                                                         | STD REF.#                                                                                     |
|---------------------------------------------------------------------------------------------------------------------------------------------------|-----------------------------------------------------------------------------------------------|
| Tune/Reschk<br>Initial Calibration Stds<br><br>CCC<br>Internal Standard/PEM<br>ICV/I.BLK<br>Surrogate Standard<br>MS/MSD Standard<br>LCS Standard | VP134642<br>VP134633,VP134634,VP134635,VP134638,VP134639,VP134640<br><br><br><br><br>VP134641 |

| Sr# | SampleID    | ClientID    | Data File Name | Date-Time         | Comment                        | Operator | Status |
|-----|-------------|-------------|----------------|-------------------|--------------------------------|----------|--------|
| 1   | BFB         | BFB         | VX046859.D     | 02 Jul 2025 11:12 |                                | JC/MD    | Ok     |
| 2   | VSTDICC001  | VSTDICC001  | VX046860.D     | 02 Jul 2025 12:11 |                                | JC/MD    | Ok,M   |
| 3   | VSTDICC005  | VSTDICC005  | VX046861.D     | 02 Jul 2025 12:37 | %D Failed for com.#13 in 5 ppb | JC/MD    | Ok,M   |
| 4   | VSTDICC020  | VSTDICC020  | VX046862.D     | 02 Jul 2025 13:18 |                                | JC/MD    | Ok,M   |
| 5   | VSTDICCC050 | VSTDICCC050 | VX046863.D     | 02 Jul 2025 13:39 | LR-13                          | JC/MD    | Ok,M   |
| 6   | VSTDICC100  | VSTDICC100  | VX046864.D     | 02 Jul 2025 14:10 |                                | JC/MD    | Ok,M   |
| 7   | VSTDICC150  | VSTDICC150  | VX046865.D     | 02 Jul 2025 14:31 |                                | JC/MD    | Ok,M   |
| 8   | VIBLK       | VIBLK       | VX046866.D     | 02 Jul 2025 14:53 |                                | JC/MD    | Ok     |
| 9   | VSTDICV050  | ICVVX070225 | VX046867.D     | 02 Jul 2025 15:14 |                                | JC/MD    | Ok,M   |

M : Manual Integration

Instrument ID: MSVOA\_X

**Daily Analysis Runlog For Sequence/QC Batch ID # VX070325**

|              |               |                   |                                   |
|--------------|---------------|-------------------|-----------------------------------|
| Review By    | John Carlone  | Review On         | 7/7/2025 9:19:20 AM               |
| Supervise By | Maresh Dadoda | Supervise On      | 7/7/2025 2:14:03 PM               |
| SubDirectory | VX070325      | HP Acquire Method | HP Processing Method 82X070225W.M |

| STD. NAME                                                                                          | STD REF.#         |
|----------------------------------------------------------------------------------------------------|-------------------|
| Tune/Reschk<br>Initial Calibration Stds                                                            | VP134621          |
| CCC<br>Internal Standard/PEM<br>ICV/I.BLK<br>Surrogate Standard<br>MS/MSD Standard<br>LCS Standard | VP134622,VP134623 |

| Sr# | SampleID     | ClientID          | Data File Name | Date-Time         | Comment        | Operator | Status |
|-----|--------------|-------------------|----------------|-------------------|----------------|----------|--------|
| 1   | BFB          | BFB               | VX046868.D     | 03 Jul 2025 07:52 |                | JC/MD    | Ok     |
| 2   | VSTDCCC050   | VSTDCCC050        | VX046869.D     | 03 Jul 2025 08:19 | pH#Lot#V12668  | JC/MD    | Ok,M   |
| 3   | VX0703MBL01  | VX0703MBL01       | VX046870.D     | 03 Jul 2025 08:42 |                | JC/MD    | Ok     |
| 4   | VX0703WBL01  | VX0703WBL01       | VX046871.D     | 03 Jul 2025 09:09 |                | JC/MD    | Ok     |
| 5   | Q2458-10     | FB-06272025       | VX046872.D     | 03 Jul 2025 09:35 | vial A pH<2 FB | JC/MD    | Ok     |
| 6   | VX0703WBS01  | VX0703WBS01       | VX046873.D     | 03 Jul 2025 09:56 |                | JC/MD    | Ok,M   |
| 7   | VX0703WBSD01 | VX0703WBSD01      | VX046874.D     | 03 Jul 2025 10:23 |                | JC/MD    | Ok,M   |
| 8   | Q2455-01     | MW1               | VX046875.D     | 03 Jul 2025 10:44 | vial A pH<2    | JC/MD    | Ok,M   |
| 9   | Q2455-03     | MW10              | VX046876.D     | 03 Jul 2025 11:04 | vial A pH<2    | JC/MD    | Ok,M   |
| 10  | Q2455-02     | MW3               | VX046877.D     | 03 Jul 2025 11:25 | vial A pH<2    | JC/MD    | Ok     |
| 11  | Q2455-04     | FIELD-BLANK       | VX046878.D     | 03 Jul 2025 11:46 | vial A pH<2 FB | JC/MD    | Ok     |
| 12  | Q2447-07ME   | LAW-25-0100ME     | VX046879.D     | 03 Jul 2025 12:07 |                | JC/MD    | Ok     |
| 13  | VX0703MBS01  | VX0703MBS01       | VX046880.D     | 03 Jul 2025 12:27 |                | JC/MD    | Ok,M   |
| 14  | IBLK         | IBLK              | VX046881.D     | 03 Jul 2025 12:53 |                | JC/MD    | Ok     |
| 15  | Q2501-01     | STORAGE-BLANK-SO  | VX046882.D     | 03 Jul 2025 13:13 | vial A pH<2    | JC/MD    | Ok     |
| 16  | Q2501-02     | STORAGE-BLANK-WA  | VX046883.D     | 03 Jul 2025 13:34 | vial A pH<2    | JC/MD    | Ok     |
| 17  | Q2501-03     | STORAGE-BLANK-WA  | VX046884.D     | 03 Jul 2025 13:55 | vial A pH<2    | JC/MD    | Ok     |
| 18  | Q2501-04     | STORAGE-BLANK-SAI | VX046885.D     | 03 Jul 2025 14:16 | vial A pH<2    | JC/MD    | Ok     |

Instrument ID: MSVOA\_X

**Daily Analysis Runlog For Sequence/QC Batch ID # VX070325**

| Review By                                                                                          | John Carlone      | Review On         | 7/7/2025 9:19:20 AM               |
|----------------------------------------------------------------------------------------------------|-------------------|-------------------|-----------------------------------|
| Supervise By                                                                                       | Maresh Dadoda     | Supervise On      | 7/7/2025 2:14:03 PM               |
| SubDirectory                                                                                       | VX070325          | HP Acquire Method | HP Processing Method 82X070225W.M |
| STD. NAME                                                                                          | STD REF.#         |                   |                                   |
| Tune/Reschk<br>Initial Calibration Stds                                                            | VP134621          |                   |                                   |
| CCC<br>Internal Standard/PEM<br>ICV/I.BLK<br>Surrogate Standard<br>MS/MSD Standard<br>LCS Standard | VP134622,VP134623 |                   |                                   |

|    |            |                  |            |                   |                              |       |      |
|----|------------|------------------|------------|-------------------|------------------------------|-------|------|
| 19 | Q2503-01   | GCAP2W           | VX046886.D | 03 Jul 2025 14:37 | vial A pH<2                  | JC/MD | Ok   |
| 20 | Q2503-02   | GCAP3W           | VX046887.D | 03 Jul 2025 14:58 | vial A pH<2                  | JC/MD | Ok   |
| 21 | Q2465-02   | 62025-B          | VX046888.D | 03 Jul 2025 15:19 | vial A pH<2 oily dark sample | JC/MD | Ok,M |
| 22 | Q2509-01   | AUD-25-0110-0111 | VX046889.D | 03 Jul 2025 15:40 | vial A pH<2 turbid sample    | JC/MD | Ok   |
| 23 | VSTDCCC050 | VSTDCCC050EC     | VX046890.D | 03 Jul 2025 16:03 |                              | JC/MD | Ok,M |

M : Manual Integration

Instrument ID: MSVOA\_Y

**Daily Analysis Runlog For Sequence/QC Batch ID # VY062325**

|              |                     |                   |                      |                      |              |
|--------------|---------------------|-------------------|----------------------|----------------------|--------------|
| Review By    | Mahesh Dadoda       | Review On         | 6/24/2025 8:24:26 AM |                      |              |
| Supervise By | Semsettin Yesilyurt | Supervise On      | 6/24/2025 8:29:32 AM |                      |              |
| SubDirectory | VY062325            | HP Acquire Method | MSVOA_Y              | HP Processing Method | 82y062325s.m |

| STD. NAME                | STD REF.#                                             |
|--------------------------|-------------------------------------------------------|
| Tune/Reschk              | VP134461                                              |
| Initial Calibration Stds | VP134462,VP134463,VP134464,VP134465,VP134466,VP134467 |
| CCC                      |                                                       |
| Internal Standard/PEM    | VP133934                                              |
| ICV/I.BLK                | VP134468                                              |
| Surrogate Standard       |                                                       |
| MS/MSD Standard          |                                                       |
| LCS Standard             |                                                       |

| Sr# | SampleID   | ClientID    | Data File Name | Date-Time         | Comment | Operator | Status |
|-----|------------|-------------|----------------|-------------------|---------|----------|--------|
| 1   | BFB        | BFB         | VY022775.D     | 23 Jun 2025 10:17 |         | SY/MD    | Ok     |
| 2   | VSTDIC005  | VSTDIC005   | VY022776.D     | 23 Jun 2025 13:38 | LR- 58  | SY/MD    | Ok,M   |
| 3   | VSTDIC010  | VSTDIC010   | VY022777.D     | 23 Jun 2025 14:00 |         | SY/MD    | Ok,M   |
| 4   | VSTDIC020  | VSTDIC020   | VY022778.D     | 23 Jun 2025 14:23 |         | SY/MD    | Ok,M   |
| 5   | VSTDIC050  | VSTDIC050   | VY022779.D     | 23 Jun 2025 14:46 |         | SY/MD    | Ok,M   |
| 6   | VSTDIC100  | VSTDIC100   | VY022780.D     | 23 Jun 2025 15:08 |         | SY/MD    | Ok,M   |
| 7   | VSTDIC150  | VSTDIC150   | VY022781.D     | 23 Jun 2025 15:31 |         | SY/MD    | Ok,M   |
| 8   | VIBLK      | VIBLK       | VY022782.D     | 23 Jun 2025 15:54 |         | SY/MD    | Ok     |
| 9   | VSTDICV050 | ICVVY062325 | VY022783.D     | 23 Jun 2025 16:17 |         | SY/MD    | Ok,M   |

M : Manual Integration

Instrument ID: MSVOA\_Y

**Daily Analysis Runlog For Sequence/QC Batch ID # VY063025**

|              |                     |                      |                      |
|--------------|---------------------|----------------------|----------------------|
| Review By    | Maresh Dadoda       | Review On            | 7/1/2025 12:22:53 PM |
| Supervise By | Semsettin Yesilyurt | Supervise On         | 7/1/2025 12:24:58 PM |
| SubDirectory | VY063025            | HP Acquire Method    | MSVOA_Y              |
|              |                     | HP Processing Method | 82y062325s.m         |

| STD. NAME                                                                                          | STD REF.#                     |
|----------------------------------------------------------------------------------------------------|-------------------------------|
| Tune/Reschk<br>Initial Calibration Stds                                                            | VP134579                      |
| CCC<br>Internal Standard/PEM<br>ICV/I.BLK<br>Surrogate Standard<br>MS/MSD Standard<br>LCS Standard | VP134588,VP134589<br>VP133934 |

| Sr# | SampleID     | ClientID     | Data File Name | Date-Time         | Comment                                 | Operator | Status   |
|-----|--------------|--------------|----------------|-------------------|-----------------------------------------|----------|----------|
| 1   | BFB          | BFB          | VY022871.D     | 30 Jun 2025 08:42 |                                         | SY/MD    | Ok       |
| 2   | VSTDCCC050   | VSTDCCC050   | VY022872.D     | 30 Jun 2025 11:03 |                                         | SY/MD    | Ok,M     |
| 3   | VY0630SBL01  | VY0630SBL01  | VY022873.D     | 30 Jun 2025 11:39 |                                         | SY/MD    | Ok       |
| 4   | VY0630SBS01  | VY0630SBS01  | VY022874.D     | 30 Jun 2025 12:09 |                                         | SY/MD    | Ok,M     |
| 5   | VY0630SBSD01 | VY0630SBSD01 | VY022875.D     | 30 Jun 2025 12:32 |                                         | SY/MD    | Ok,M     |
| 6   | Q2452-07     | EP-2-VOC     | VY022876.D     | 30 Jun 2025 13:08 | vial-B                                  | SY/MD    | Ok,M     |
| 7   | Q2436-03     | TP-85        | VY022877.D     | 30 Jun 2025 13:32 | vial-B                                  | SY/MD    | Ok       |
| 8   | Q2436-04     | TP-86        | VY022878.D     | 30 Jun 2025 14:28 | vial-B                                  | SY/MD    | Ok       |
| 9   | Q2436-05     | TP-84        | VY022879.D     | 30 Jun 2025 14:52 | vial-B                                  | SY/MD    | Ok       |
| 10  | Q2457-01     | OK-2-062725  | VY022880.D     | 30 Jun 2025 15:15 | vial-A Internal Standard Fail           | SY/MD    | ReRun    |
| 11  | Q2450-01     | PL-2-062725  | VY022881.D     | 30 Jun 2025 15:38 | vial-A                                  | SY/MD    | Ok       |
| 12  | Q2449-01     | RBR200032    | VY022882.D     | 30 Jun 2025 16:02 | vial-A                                  | SY/MD    | Ok       |
| 13  | Q2447-02     | COMP-1       | VY022883.D     | 30 Jun 2025 16:25 | vial-A                                  | SY/MD    | Ok       |
| 14  | Q2447-05     | COMP-2       | VY022884.D     | 30 Jun 2025 16:49 | vial-A Not Purge                        | SY/MD    | Not Ok   |
| 15  | Q2447-07     | LAW-25-0100  | VY022885.D     | 30 Jun 2025 17:12 | vial-A Internal standard fail;Need MeOH | SY/MD    | Dilution |
| 16  | Q2459-03     | TP-1-VOC     | VY022886.D     | 30 Jun 2025 17:35 | vial-A                                  | SY/MD    | Ok       |
| 17  | Q2458-01     | TP-76        | VY022887.D     | 30 Jun 2025 17:59 | vial-A                                  | SY/MD    | Ok       |

Instrument ID: MSVOA\_Y

**Daily Analysis Runlog For Sequence/QC Batch ID # VY063025**

|              |                     |                   |                      |                      |              |
|--------------|---------------------|-------------------|----------------------|----------------------|--------------|
| Review By    | Mahesh Dadoda       | Review On         | 7/1/2025 12:22:53 PM |                      |              |
| Supervise By | Semsettin Yesilyurt | Supervise On      | 7/1/2025 12:24:58 PM |                      |              |
| SubDirectory | VY063025            | HP Acquire Method | MSVOA_Y              | HP Processing Method | 82y062325s.m |

| STD. NAME                               | STD REF.#         |
|-----------------------------------------|-------------------|
| Tune/Reschk<br>Initial Calibration Stds | VP134579          |
| CCC                                     | VP134588,VP134589 |
| Internal Standard/PEM                   | VP133934          |
| ICV/I.BLK                               |                   |
| Surrogate Standard                      |                   |
| MS/MSD Standard                         |                   |
| LCS Standard                            |                   |

|    |          |       |            |                   |                               |       |       |
|----|----------|-------|------------|-------------------|-------------------------------|-------|-------|
| 18 | Q2458-02 | TP-55 | VY022888.D | 30 Jun 2025 18:22 | vial-A                        | SY/MD | Ok    |
| 19 | Q2458-03 | TP-68 | VY022889.D | 30 Jun 2025 18:46 | vial-A Internal Standard Fail | SY/MD | ReRun |

M : Manual Integration

Instrument ID: MSVOA\_Y

**Daily Analysis Runlog For Sequence/QC Batch ID # VY070125**

|              |                     |                   |                      |                      |              |
|--------------|---------------------|-------------------|----------------------|----------------------|--------------|
| Review By    | Mahesh Dadoda       | Review On         | 7/2/2025 12:29:56 PM |                      |              |
| Supervise By | Semsettin Yesilyurt | Supervise On      | 7/2/2025 12:30:51 PM |                      |              |
| SubDirectory | VY070125            | HP Acquire Method | MSVOA_Y              | HP Processing Method | 82y062325s.m |

| STD. NAME                                                                                          | STD REF.#                     |
|----------------------------------------------------------------------------------------------------|-------------------------------|
| Tune/Reschk<br>Initial Calibration Stds                                                            | VP134591                      |
| CCC<br>Internal Standard/PEM<br>ICV/I.BLK<br>Surrogate Standard<br>MS/MSD Standard<br>LCS Standard | VP134592,VP134593<br>VP133934 |

| Sr# | SampleID     | ClientID      | Data File Name | Date-Time         | Comment                       | Operator | Status   |
|-----|--------------|---------------|----------------|-------------------|-------------------------------|----------|----------|
| 1   | BFB          | BFB           | VY022890.D     | 01 Jul 2025 08:47 |                               | SY/MD    | Ok       |
| 2   | VSTDCCC050   | VSTDCCC050    | VY022891.D     | 01 Jul 2025 10:32 |                               | SY/MD    | Ok,M     |
| 3   | VY0701SBL01  | VY0701SBL01   | VY022892.D     | 01 Jul 2025 11:05 |                               | SY/MD    | Ok       |
| 4   | VY0701SBS01  | VY0701SBS01   | VY022893.D     | 01 Jul 2025 11:34 |                               | SY/MD    | Ok,M     |
| 5   | VY0701SBSD01 | VY0701SBSD01  | VY022894.D     | 01 Jul 2025 11:56 |                               | SY/MD    | Ok,M     |
| 6   | Q2457-01RE   | OK-2-062725RE | VY022895.D     | 01 Jul 2025 12:34 | vial-B Internal Standard Fail | SY/MD    | Confirms |
| 7   | Q2447-05     | COMP-2        | VY022896.D     | 01 Jul 2025 12:57 | vial-B Internal Standard Fail | SY/MD    | Ok       |
| 8   | Q2458-03     | TP-68         | VY022897.D     | 01 Jul 2025 13:20 | vial-B                        | SY/MD    | Ok       |
| 9   | Q2458-04     | TP-67         | VY022898.D     | 01 Jul 2025 14:23 | vial-A Internal Standard Fail | SY/MD    | ReRun    |
| 10  | Q2458-05     | TP-66         | VY022899.D     | 01 Jul 2025 15:06 | vial-A                        | SY/MD    | Ok       |
| 11  | VIBLK        | VIBLK         | VY022900.D     | 01 Jul 2025 15:41 |                               | SY/MD    | Ok       |
| 12  | VSTDCCC050   | VSTDCCC050EC  | VY022901.D     | 01 Jul 2025 16:27 |                               | SY/MD    | Ok,M     |

M : Manual Integration

Instrument ID: MSVOA\_Y

**Daily Analysis Runlog For Sequence/QC Batch ID # VY070225**

|              |                     |                   |                                           |
|--------------|---------------------|-------------------|-------------------------------------------|
| Review By    | Maresh Dadoda       | Review On         | 7/3/2025 3:08:26 PM                       |
| Supervise By | Semsettin Yesilyurt | Supervise On      | 7/3/2025 3:13:05 PM                       |
| SubDirectory | VY070225            | HP Acquire Method | MSVOA_Y HP Processing Method 82y062325s.m |

| STD. NAME                                                                                          | STD REF.#                     |
|----------------------------------------------------------------------------------------------------|-------------------------------|
| Tune/Reschk<br>Initial Calibration Stds                                                            | VP134605                      |
| CCC<br>Internal Standard/PEM<br>ICV/I.BLK<br>Surrogate Standard<br>MS/MSD Standard<br>LCS Standard | VP134608,VP134609<br>VP133934 |

| Sr# | SampleID     | ClientID     | Data File Name | Date-Time         | Comment                       | Operator | Status |
|-----|--------------|--------------|----------------|-------------------|-------------------------------|----------|--------|
| 1   | BFB          | BFB          | VY022902.D     | 02 Jul 2025 08:51 |                               | SY/MD    | Ok     |
| 2   | VSTDCCC050   | VSTDCCC050   | VY022903.D     | 02 Jul 2025 10:37 | CCAL failed high for comp.#16 | SY/MD    | Ok,M   |
| 3   | VY0702SBL01  | VY0702SBL01  | VY022904.D     | 02 Jul 2025 11:42 |                               | SY/MD    | Ok     |
| 4   | VY0702SBS01  | VY0702SBS01  | VY022905.D     | 02 Jul 2025 12:12 |                               | SY/MD    | Ok,M   |
| 5   | VY0702SBSD01 | VY0702SBSD01 | VY022906.D     | 02 Jul 2025 12:35 |                               | SY/MD    | Ok,M   |
| 6   | Q2469-03     | WC-1-VOC     | VY022907.D     | 02 Jul 2025 13:12 | vial-A                        | SY/MD    | Ok     |
| 7   | Q2464-01     | OR-3-063025  | VY022908.D     | 02 Jul 2025 13:36 | vial-A                        | SY/MD    | Ok     |
| 8   | Q2478-03     | WC-1-VOC     | VY022909.D     | 02 Jul 2025 13:59 | vial-A                        | SY/MD    | Ok     |
| 9   | Q2475-01     | SOIL-PILE    | VY022910.D     | 02 Jul 2025 14:23 | vial-A                        | SY/MD    | Ok     |
| 10  | Q2486-02     | VOC          | VY022911.D     | 02 Jul 2025 14:46 | vial-A Internal Standard Fail | SY/MD    | ReRun  |
| 11  | Q2458-04     | TP-67        | VY022912.D     | 02 Jul 2025 15:09 | vial-B                        | SY/MD    | Ok     |
| 12  | Q2458-06     | TP-60        | VY022913.D     | 02 Jul 2025 15:33 | vial-A                        | SY/MD    | Ok     |
| 13  | Q2458-07     | TP-62        | VY022914.D     | 02 Jul 2025 15:56 | vial-A                        | SY/MD    | Ok     |
| 14  | Q2458-08     | TP-63        | VY022915.D     | 02 Jul 2025 16:20 | vial-A                        | SY/MD    | Ok     |
| 15  | Q2458-09     | TP-59        | VY022916.D     | 02 Jul 2025 16:43 | vial-A CCC Failed High        | SY/MD    | ReRun  |
| 16  | Q2473-01     | PIT#1        | VY022917.D     | 02 Jul 2025 17:07 | vial-A                        | SY/MD    | Ok     |
| 17  | Q2473-02     | PIT#2        | VY022918.D     | 02 Jul 2025 17:30 | vial-A                        | SY/MD    | Ok     |
| 18  | Q2473-03     | PIT#3        | VY022919.D     | 02 Jul 2025 17:53 | vial-A                        | SY/MD    | Ok     |



Instrument ID: MSVOA\_Y

**Daily Analysis Runlog For Sequence/QCBatch ID # VY070225**

| Review By                               | Mahesh Dadoda       | Review On         | 7/3/2025 3:08:26 PM |                      |              |
|-----------------------------------------|---------------------|-------------------|---------------------|----------------------|--------------|
| Supervise By                            | Semsettin Yesilyurt | Supervise On      | 7/3/2025 3:13:05 PM |                      |              |
| SubDirectory                            | VY070225            | HP Acquire Method | MSVOA_Y             | HP Processing Method | 82y062325s.m |
| STD. NAME                               |                     | STD REF.#         |                     |                      |              |
| Tune/Reschk<br>Initial Calibration Stds |                     | VP134605          |                     |                      |              |
| CCC                                     |                     | VP134608,VP134609 |                     |                      |              |
| Internal Standard/PEM                   |                     | VP133934          |                     |                      |              |
| ICV/I.BLK                               |                     |                   |                     |                      |              |
| Surrogate Standard                      |                     |                   |                     |                      |              |
| MS/MSD Standard                         |                     |                   |                     |                      |              |
| LCS Standard                            |                     |                   |                     |                      |              |

|    |            |              |            |                   |                                              |       |        |
|----|------------|--------------|------------|-------------------|----------------------------------------------|-------|--------|
| 19 | Q2473-04   | PIT#4        | VY022920.D | 02 Jul 2025 18:17 | vial-A Surrogate fail;CCC Failed High        | SY/MD | ReRun  |
| 20 | Q2487-01   | G4(1.5)      | VY022921.D | 02 Jul 2025 18:40 | vial-A CCC failed high                       | SY/MD | ReRun  |
| 21 | Q2487-02   | G4(10)       | VY022922.D | 02 Jul 2025 19:04 | vial-A CCC failed high                       | SY/MD | ReRun  |
| 22 | Q2487-03   | G3(9)        | VY022923.D | 02 Jul 2025 19:27 | vial-A CCC failed high                       | SY/MD | ReRun  |
| 23 | Q2487-04   | G3(3)        | VY022924.D | 02 Jul 2025 19:50 | vial-A Internal standard fail;Surrogate fail | SY/MD | ReRun  |
| 24 | Q2487-05   | G2(2.5)      | VY022925.D | 02 Jul 2025 20:14 | vial-A Internal Standard Fail                | SY/MD | ReRun  |
| 25 | VIBLK      | VIBLK        | VY022926.D | 02 Jul 2025 20:37 |                                              | SY/MD | Ok     |
| 26 | VSTDCCC050 | VSTDCCC050EC | VY022927.D | 02 Jul 2025 21:23 | Out of Tune                                  | SY/MD | Not Ok |

M : Manual Integration

Instrument ID: MSVOA\_Y

**Daily Analysis Runlog For Sequence/QC Batch ID # VY070325**

|              |                     |                   |                                           |
|--------------|---------------------|-------------------|-------------------------------------------|
| Review By    | Mahesh Dadoda       | Review On         | 7/7/2025 2:11:58 PM                       |
| Supervise By | Semsettin Yesilyurt | Supervise On      | 7/7/2025 2:15:30 PM                       |
| SubDirectory | VY070325            | HP Acquire Method | MSVOA_Y HP Processing Method 82y062325s.m |

| STD. NAME                                                                                          | STD REF.#                     |
|----------------------------------------------------------------------------------------------------|-------------------------------|
| Tune/Reschk<br>Initial Calibration Stds                                                            | VP134628                      |
| CCC<br>Internal Standard/PEM<br>ICV/I.BLK<br>Surrogate Standard<br>MS/MSD Standard<br>LCS Standard | VP134631,VP134632<br>VP133934 |

| Sr# | SampleID     | ClientID     | Data File Name | Date-Time         | Comment                                          | Operator | Status   |
|-----|--------------|--------------|----------------|-------------------|--------------------------------------------------|----------|----------|
| 1   | BFB          | BFB          | VY022928.D     | 03 Jul 2025 08:03 |                                                  | SY/MD    | Ok       |
| 2   | VSTDCCC050   | VSTDCCC050   | VY022929.D     | 03 Jul 2025 09:26 |                                                  | SY/MD    | Ok,M     |
| 3   | VY0703SBL01  | VY0703SBL01  | VY022930.D     | 03 Jul 2025 09:57 |                                                  | SY/MD    | Ok       |
| 4   | VY0703SBS01  | VY0703SBS01  | VY022931.D     | 03 Jul 2025 10:29 |                                                  | SY/MD    | Ok,M     |
| 5   | VY0703SBSD01 | VY0703SBSD01 | VY022932.D     | 03 Jul 2025 10:52 |                                                  | SY/MD    | Ok,M     |
| 6   | Q2493-03     | WC-11-VOC    | VY022933.D     | 03 Jul 2025 11:29 | vial-A                                           | SY/MD    | Ok       |
| 7   | Q2491-01     | EO-1-070225  | VY022934.D     | 03 Jul 2025 11:53 | vial-A                                           | SY/MD    | Ok       |
| 8   | Q2486-02     | VOC          | VY022935.D     | 03 Jul 2025 12:16 | vial-B                                           | SY/MD    | Ok       |
| 9   | Q2473-04RE   | PIT#4RE      | VY022936.D     | 03 Jul 2025 12:39 | vial-B Internal Standard Fail;<br>Surrogate Fail | SY/MD    | Confirms |
| 10  | Q2458-09     | TP-59        | VY022937.D     | 03 Jul 2025 13:03 | vial-B                                           | SY/MD    | Ok       |
| 11  | Q2487-01     | G4(1.5)      | VY022938.D     | 03 Jul 2025 13:26 | vial-B                                           | SY/MD    | Ok       |
| 12  | Q2487-02     | G4(10)       | VY022939.D     | 03 Jul 2025 13:50 | vial-B                                           | SY/MD    | Ok       |
| 13  | Q2487-03     | G3(9)        | VY022940.D     | 03 Jul 2025 14:13 | vial-B                                           | SY/MD    | Ok       |
| 14  | Q2487-04     | G3(3)        | VY022941.D     | 03 Jul 2025 14:36 | vial-B                                           | SY/MD    | Ok       |
| 15  | Q2487-05RE   | G2(2.5)RE    | VY022942.D     | 03 Jul 2025 15:00 | vial-B Internal Standard Fail                    | SY/MD    | Confirms |
| 16  | Q2487-06     | G2(9)        | VY022943.D     | 03 Jul 2025 15:23 | vial-A                                           | SY/MD    | Ok       |
| 17  | Q2487-07     | G1(4.5)      | VY022944.D     | 03 Jul 2025 15:47 | vial-A Internal Standard Fail;<br>Surrogate Fail | SY/MD    | ReRun    |

M : Manual Integration