

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: CAMP02
 Lab Code: ACE SDG NO.: Q2458
 Lab File ID: VX046869.D Date Analyzed: 07/03/2025
 Instrument ID: MSVOA_X Time Analyzed: 08:19
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	416385	5.56	663948	6.76	582621	10.05
UPPER LIMIT	832770	6.056	1327900	7.263	1165240	10.549
LOWER LIMIT	208193	5.056	331974	6.263	291311	9.549
EPA SAMPLE NO.						
FB-06272025	397171	5.56	689071	6.76	628783	10.05
VX0703WBL01	435949	5.56	752530	6.76	691838	10.05
VX0703WBS01	380935	5.56	631809	6.77	572542	10.06
VX0703WBSD01	339915	5.56	568093	6.76	517645	10.05

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.