

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: CAMP02  
 Lab Code: ACE SDG NO.: Q2458  
 Lab File ID: VY022872.D Date Analyzed: 06/30/2025  
 Instrument ID: MSVOA\_Y Time Analyzed: 11:03  
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

|                | IS1<br>AREA # | RT #  | IS2<br>AREA # | RT #  | IS3<br>AREA # | RT #   |
|----------------|---------------|-------|---------------|-------|---------------|--------|
| 12 HOUR STD    | 450935        | 7.71  | 753538        | 8.62  | 651538        | 11.41  |
| UPPER LIMIT    | 901870        | 8.207 | 1507080       | 9.115 | 1303080       | 11.914 |
| LOWER LIMIT    | 225468        | 7.207 | 376769        | 8.115 | 325769        | 10.914 |
| EPA SAMPLE NO. |               |       |               |       |               |        |
| TP-76          | 303882        | 7.71  | 573423        | 8.62  | 579838        | 11.41  |
| TP-55          | 302919        | 7.71  | 573452        | 8.62  | 562005        | 11.41  |
| VY0630SBL01    | 371114        | 7.71  | 672085        | 8.62  | 635637        | 11.41  |
| VY0630SBS01    | 473861        | 7.71  | 796065        | 8.62  | 673720        | 11.41  |
| VY0630SBSD01   | 457956        | 7.71  | 769619        | 8.62  | 654092        | 11.41  |

IS1 = Pentafluorobenzene  
 IS2 = 1,4-Difluorobenzene  
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area  
 AREA LOWER LIMIT = -50% of internal standard area  
 RT UPPER LIMIT = +0.50 minutes of internal standard RT  
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.  
 \* Values outside of QC limits.