

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: CAMP02
 Lab Code: ACE SDG NO.: Q2458
 Lab File ID: VY022891.D Date Analyzed: 07/01/2025
 Instrument ID: MSVOA_Y Time Analyzed: 10:32
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	433923	7.71	723740	8.62	623953	11.41
UPPER LIMIT	867846	8.213	1447480	9.115	1247910	11.914
LOWER LIMIT	216962	7.213	361870	8.115	311977	10.914
EPA SAMPLE NO.						
TP-68	309337	7.71	580195	8.61	586876	11.41
TP-66	227063	7.71	478916	8.62	482244	11.41
VY0701SBL01	323968	7.71	617486	8.62	600262	11.41
VY0701SBS01	427188	7.71	723243	8.62	617191	11.41

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.