

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: CAMP02
 Lab Code: ACE SDG NO.: Q2458
 Lab File ID: VY022903.D Date Analyzed: 07/02/2025
 Instrument ID: MSVOA_Y Time Analyzed: 10:37
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	453239	7.71	750175	8.62	644334	11.41
UPPER LIMIT	906478	8.213	1500350	9.116	1288670	11.914
LOWER LIMIT	226620	7.213	375088	8.116	322167	10.914
EPA SAMPLE NO.						
TP-67	314085	7.71	602578	8.62	586720	11.41
TP-60	284075	7.71	537745	8.62	538792	11.41
TP-62	277237	7.71	536046	8.62	553104	11.41
TP-63	278456	7.71	537398	8.62	515031	11.41
VY0702SBL01	327303	7.71	608180	8.62	599654	11.42
VY0702SBS01	424569	7.71	714057	8.62	598160	11.41

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.