

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: CAMP02
 Lab Code: ACE SDG NO.: Q2458
 Lab File ID: VY022929.D Date Analyzed: 07/03/2025
 Instrument ID: MSVOA_Y Time Analyzed: 09:26
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	402903	7.71	653588	8.62	556811	11.41
UPPER LIMIT	805806	8.207	1307180	9.116	1113620	11.914
LOWER LIMIT	201452	7.207	326794	8.116	278406	10.914
EPA SAMPLE NO.						
TP-59	268585	7.71	499539	8.62	495456	11.41
VY0703SBL01	297416	7.71	558302	8.62	553906	11.41
VY0703SBS01	387913	7.71	648160	8.62	537972	11.41

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.