

Initial and Continuing Calibration Blank Summary

Client: ATG-GREENVILLE AEC

SDG No.: Q2460

Project: AEC-2025-0013-CSC 4031 Watervliet

RunNo.: LB136335

Analyte		Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID:	ICB							
COD		mg/L	< 5.0000	5.0000	U	1.50	10	05/28/2025
Sample ID:	CCB1							
COD		mg/L	< 5.0000	5.0000	U	1.50	10	07/01/2025
Sample ID:	CCB2							
COD		mg/L	< 5.0000	5.0000	U	1.50	10	07/01/2025