

Matrix Spike Summary

Client:	ATG-GREENVILLE AEC	SDG No.:	Q2460
Project:	AEC-2025-0013-CSC 4031 Watervliet	Sample ID:	Q2441-01
Client ID:	MH-6272025MS	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Oil and Grease	mg/L	78-114	668		648		20.0	1	100		07/02/2025

Matrix Spike Summary

Client:	ATG-GREENVILLE AEC	SDG No.:	Q2460
Project:	AEC-2025-0013-CSC 4031 Watervliet	Sample ID:	Q2441-01
Client ID:	MH-6272025MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Oil and Grease	mg/L	78-114	668		648		20.0	1	101		07/02/2025

Matrix Spike Summary

Client:	ATG-GREENVILLE AEC	SDG No.:	Q2460
Project:	AEC-2025-0013-CSC 4031 Watervliet	Sample ID:	Q2461-01
Client ID:	SW-3MS	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
COD	mg/L	75-125	285		240	D	50.0	5	90		07/01/2025

Matrix Spike Summary

Client:	ATG-GREENVILLE AEC	SDG No.:	Q2460
Project:	AEC-2025-0013-CSC 4031 Watervliet	Sample ID:	Q2461-01
Client ID:	SW-3MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
COD	mg/L	75-125	280		240	D	50.0	5	80		07/01/2025