

Duplicate Sample Summary

Client:	ATG-GREENVILLE AEC	SDG No.:	Q2460
Project:	AEC-2025-0013-CSC 4031 Watervliet	Sample ID:	Q2441-01
Client ID:	MH-6272025MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Oil and Grease	mg/L	+/-18	668		668		1	0.03		07/02/2025

Duplicate Sample Summary

Client: ATG-GREENVILLE AEC	SDG No.: Q2460
Project: AEC-2025-0013-CSC 4031 Watervliet	Sample ID: Q2461-01
Client ID: SW-3DUP	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
COD	mg/L	+/-20	240	D	245		5	2.06		07/01/2025

Duplicate Sample Summary

Client:	ATG-GREENVILLE AEC	SDG No.:	Q2460
Project:	AEC-2025-0013-CSC 4031 Watervliet	Sample ID:	Q2461-01
Client ID:	SW-3MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
COD	mg/L	+/-20	285		280		5	1.77		07/01/2025