

Laboratory Control Sample Summary

Client: ATG-GREENVILLE AEC

SDG No.: Q2460

Project: AEC-2025-0013-CSC 4031 Watervliet

Run No.: LB136335

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136335BS							
COD	mg/L	50	48.9		98	1	90-110	07/01/2025

Laboratory Control Sample Summary

Client: ATG-GREENVILLE AEC

SDG No.: Q2460

Project: AEC-2025-0013-CSC 4031 Watervliet

Run No.: LB136347

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136347BS							
Oil and Grease	mg/L	20.0	16.9		84	1	78-114	07/02/2025