

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: ATGG01

Lab Code: CHEM Case No.: Q2460 SAS No.: Q2460 SDG No.: Q2460

Instrument ID: MSVOA_X Calibration Date/Time: 07/01/2025 08:53

Lab File ID: VX046844.D Init. Calib. Date(s): 06/17/2025 06/17/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:19 17:18

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Benzene	1.413	1.367		-3.26	20
Toluene	0.876	0.844		-3.65	20
Ethyl Benzene	1.929	1.860		-3.58	20
m/p-Xylenes	0.719	0.699		-2.78	20
o-Xylene	0.673	0.666		-1.04	20
1,2-Dichloroethane-d4	0.692	0.699		1.01	20
Dibromofluoromethane	0.331	0.380		14.8	20
Toluene-d8	1.189	1.305		9.76	20
4-Bromofluorobenzene	0.443	0.489		10.38	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.