

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance

Contract: ATGG01

Lab Code: ACE

SDG NO.: Q2460

Lab File ID: VX046843.D

BFB Injection Date: 07/01/2025

Instrument ID: MSVOA_X

BFB Injection Time: 08:11

GC Column: DB-624UI ID: 0.18 (mm)

Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	18.6
75	30.0 - 60.0% of mass 95	50.7
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.3
173	Less than 2.0% of mass 174	0.8 (1.1) 1
174	50.0 - 100.0% of mass 95	77
175	5.0 - 9.0% of mass 174	5.8 (7.5) 1
176	95.0 - 101.0% of mass 174	74.4 (96.6) 1
177	5.0 - 9.0% of mass 176	4.7 (6.3) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VX046844.D	07/01/2025	08:53
VX0701WBL01	VX0701WBL01	VX046846.D	07/01/2025	09:41
VX0701WBS01	VX0701WBS01	VX046847.D	07/01/2025	10:15
VX0701WBSD01	VX0701WBSD01	VX046848.D	07/01/2025	10:42
SW-1	Q2460-01	VX046850.D	07/01/2025	11:23