

Initial and Continuing Calibration Verification

Client: ATG-GREENVILLE AEC

SDG No.: Q2461

Project: AEC-2025-0013-CSC 4086Shenectady

RunNo.: LB136335

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: COD	ICV	mg/L	50.962	50	102	95-105	05/28/2025
Sample ID: COD	CCV1	mg/L	50.962	50	102	95-105	07/01/2025
Sample ID: COD	CCV2	mg/L	49.946	50	100	95-105	07/01/2025