

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX061725

Review By	Semsettin Yesilyurt	Review On	6/18/2025 5:28:16 PM
Supervise By	Mahesh Dadoda	Supervise On	6/18/2025 6:22:48 PM
SubDirectory	VX061725	HP Acquire Method	HP Processing Method 82X061725W.M

STD. NAME	STD REF.#
Tune/Reschk	VP134389
Initial Calibration Stds	VP134390,VP134391,VP134392,VP134393,VP134394,VP134395
CCC	
Internal Standard/PEM	
ICV/I.BLK	VP134396
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX046715.D	17 Jun 2025 08:46		JC/MD	Ok
2	VSTDICC001	VSTDICC001	VX046716.D	17 Jun 2025 10:09	RRF check	JC/MD	Not Ok
3	VSTDICC001	VSTDICC001	VX046717.D	17 Jun 2025 10:58	low response	JC/MD	ReRun
4	VSTDICC005	VSTDICC005	VX046718.D	17 Jun 2025 11:19		JC/MD	Ok,M
5	VSTDICC020	VSTDICC020	VX046719.D	17 Jun 2025 13:59	LR- 10,49	JC/MD	Ok,M
6	VSTDICCC050	VSTDICCC050	VX046720.D	17 Jun 2025 14:20		JC/MD	Ok,M
7	VSTDICC100	VSTDICC100	VX046721.D	17 Jun 2025 14:41		JC/MD	Ok,M
8	VSTDICC150	VSTDICC150	VX046722.D	17 Jun 2025 15:02		JC/MD	Ok,M
9	IBLK	IBLK	VX046723.D	17 Jun 2025 15:23		JC/MD	Ok
10	VSTDICC001	VSTDICC001	VX046724.D	17 Jun 2025 16:20	spike error	JC/MD	Not Ok
11	VSTDICC001	VSTDICC001	VX046725.D	17 Jun 2025 17:18		JC/MD	Ok,M
12	VSTDICV050	ICVVX061725	VX046726.D	17 Jun 2025 18:00		JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX070125

Review By	John Carlone	Review On	7/3/2025 8:46:34 AM
Supervise By	Mahesh Dadoda	Supervise On	7/3/2025 3:08:04 PM
SubDirectory	VX070125	HP Acquire Method	HP Processing Method 82X061725W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP134598
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134599,VP134600

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX046843.D	01 Jul 2025 08:11		JC/MD	Ok
2	VSTDCCC050	VSTDCCC050	VX046844.D	01 Jul 2025 08:53	pH#Lot#V12668	JC/MD	Ok,M
3	VX0701MBL01	VX0701MBL01	VX046845.D	01 Jul 2025 09:20		JC/MD	Ok
4	VX0701WBL01	VX0701WBL01	VX046846.D	01 Jul 2025 09:41		JC/MD	Ok
5	VX0701WBS01	VX0701WBS01	VX046847.D	01 Jul 2025 10:15		JC/MD	Ok,M
6	VX0701WBSD01	VX0701WBSD01	VX046848.D	01 Jul 2025 10:42		JC/MD	Ok,M
7	Q2461-01	SW-3	VX046849.D	01 Jul 2025 11:03	vial A pH<2	JC/MD	Ok
8	Q2460-01	SW-1	VX046850.D	01 Jul 2025 11:23	vial A pH<2 yellow foamy sample	JC/MD	Ok
9	Q2463-01	TW-WTS-11	VX046851.D	01 Jul 2025 11:54	vial A pH<2	JC/MD	Ok
10	IBLK	IBLK	VX046852.D	01 Jul 2025 12:15		JC/MD	Ok
11	Q2455-01	MW1	VX046853.D	01 Jul 2025 12:36	CCC failing Low for com#11, Need lower dilution	JC/MD	Not Ok
12	Q2455-02	MW3	VX046854.D	01 Jul 2025 12:57	CCC failing Low for com#11	JC/MD	Not Ok
13	Q2455-03	MW10	VX046855.D	01 Jul 2025 13:18	CCC failing Low for com#11, Need lower dilution	JC/MD	Not Ok
14	IBLK	IBLK	VX046856.D	01 Jul 2025 13:39		JC/MD	Ok
15	IBLK	IBLK	VX046857.D	01 Jul 2025 14:00		JC/MD	Ok
16	VSTDCCC050	VSTDCCC050EC	VX046858.D	01 Jul 2025 16:04		JC/MD	Ok,M

M : Manual Integration