

Surrogate Summary

SDG No.: Q2461

Client: ATG-GREENVILLE AEC

Analytical Method: SW8260-Low

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q2461-01	SW-3	1,2-Dichloroethane-d4	50	47.7	95		74	125
		Dibromofluoromethane	50	49.1	98		75	124
		Toluene-d8	50	49.0	98		86	113
		4-Bromofluorobenzene	50	51.6	103		77	121
VX0701WBL01	VX0701WBL01	1,2-Dichloroethane-d4	50	49.7	99		74	125
		Dibromofluoromethane	50	51.1	102		75	124
		Toluene-d8	50	50.5	101		86	113
		4-Bromofluorobenzene	50	52.2	104		77	121
VX0701WBS01	VX0701WBS01	1,2-Dichloroethane-d4	50	42.7	85		74	125
		Dibromofluoromethane	50	46.8	94		75	124
		Toluene-d8	50	47.3	94		86	113
		4-Bromofluorobenzene	50	48.0	96		77	121
VX0701WBSD0	VX0701WBSD01	1,2-Dichloroethane-d4	50	45.2	90		74	125
		Dibromofluoromethane	50	48.3	97		75	124
		Toluene-d8	50	47.9	96		86	113
		4-Bromofluorobenzene	50	49.0	98		77	121