

Duplicate Sample Summary

Client:	ENTACT	SDG No.:	Q2463
Project:	540 Degraw St, Brooklyn, NY - E9309	Sample ID:	Q2462-01
Client ID:	40425DUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Flash Point	o F	+/-2	>212.0		>212.0		1	0		07/01/2025

Duplicate Sample Summary

Client:	ENTACT	SDG No.:	Q2463
Project:	540 Degraw St, Brooklyn, NY - E9309	Sample ID:	Q2463-01
Client ID:	TW-WTS-11DUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
TSS	mg/L	+/-5	120		120		1	0.25		07/02/2025

Duplicate Sample Summary

Client:	ENTACT	SDG No.:	Q2463
Project:	540 Degraw St, Brooklyn, NY - E9309	Sample ID:	Q2499-02
Client ID:	CompDUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
BOD5	mg/L	+/-20	726		708		1	2.51		07/02/2025