



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: ENTACT

SDG No.: Q2463

Contract: ENTA05

Lab Code: ACE

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.22	0.2	112	70 - 130	CV	07/01/2025	11:52	LB136337
CRI01	Cadmium	6.26	6.0	104	65 - 135	P	07/03/2025	12:26	LB136371
	Copper	21.0	20.0	105	65 - 135	P	07/03/2025	12:26	LB136371
	Lead	12.8	12.0	106	65 - 135	P	07/03/2025	12:26	LB136371
	Nickel	40.5	40.0	101	65 - 135	P	07/03/2025	12:26	LB136371
	Zinc	44.9	40.0	112	65 - 135	P	07/03/2025	12:26	LB136371