



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ENTACT

SDG No.: Q2463

Contract: ENTA05

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB13	Mercury	0.076	+/-0.2	U	0.20	CV	07/01/2025	11:45	LB136337

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ENTACT

SDG No.: Q2463

Contract: ENTA05

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB40	Mercury	0.076	+/-0.2	U	0.20	CV	07/01/2025	11:49	LB136337
CCB41	Mercury	0.076	+/-0.2	U	0.20	CV	07/01/2025	12:19	LB136337
CCB42	Mercury	0.076	+/-0.2	U	0.20	CV	07/01/2025	12:49	LB136337
CCB43	Mercury	0.076	+/-0.2	U	0.20	CV	07/01/2025	13:13	LB136337

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ENTACT

SDG No.: Q2463

Contract: ENTA05

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Cadmium	0.50	+/-3	U	6.00	P	07/03/2025	12:21	LB136371
	Copper	4.60	+/-10	U	20.0	P	07/03/2025	12:21	LB136371
	Lead	2.30	+/-6	U	12.0	P	07/03/2025	12:21	LB136371
	Nickel	3.06	+/-20	U	40.0	P	07/03/2025	12:21	LB136371
	Zinc	16.7	+/-20	U	40.0	P	07/03/2025	12:21	LB136371

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ENTACT

SDG No.: Q2463

Contract: ENTA05

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Cadmium	0.50	+/-3	U	6.00	P	07/03/2025	13:27	LB136371
	Copper	4.60	+/-10	U	20.0	P	07/03/2025	13:27	LB136371
	Lead	2.30	+/-6	U	12.0	P	07/03/2025	13:27	LB136371
	Nickel	3.06	+/-20	U	40.0	P	07/03/2025	13:27	LB136371
	Zinc	16.7	+/-20	U	40.0	P	07/03/2025	13:27	LB136371
CCB02	Cadmium	0.50	+/-3	U	6.00	P	07/03/2025	14:55	LB136371
	Copper	4.60	+/-10	U	20.0	P	07/03/2025	14:55	LB136371
	Lead	2.30	+/-6	U	12.0	P	07/03/2025	14:55	LB136371
	Nickel	3.06	+/-20	U	40.0	P	07/03/2025	14:55	LB136371
	Zinc	16.7	+/-20	U	40.0	P	07/03/2025	14:55	LB136371
CCB03	Cadmium	0.50	+/-3	U	6.00	P	07/03/2025	16:08	LB136371
	Copper	4.60	+/-10	U	20.0	P	07/03/2025	16:08	LB136371
	Lead	2.30	+/-6	U	12.0	P	07/03/2025	16:08	LB136371
	Nickel	3.06	+/-20	U	40.0	P	07/03/2025	16:08	LB136371
	Zinc	16.7	+/-20	U	40.0	P	07/03/2025	16:08	LB136371
CCB04	Cadmium	0.50	+/-3	U	6.00	P	07/03/2025	17:02	LB136371
	Copper	4.60	+/-10	U	20.0	P	07/03/2025	17:02	LB136371
	Lead	2.30	+/-6	U	12.0	P	07/03/2025	17:02	LB136371
	Nickel	3.06	+/-20	U	40.0	P	07/03/2025	17:02	LB136371
	Zinc	16.7	+/-20	U	40.0	P	07/03/2025	17:02	LB136371
CCB05	Cadmium	0.50	+/-3	U	6.00	P	07/03/2025	17:32	LB136371
	Copper	4.60	+/-10	U	20.0	P	07/03/2025	17:32	LB136371
	Lead	2.30	+/-6	U	12.0	P	07/03/2025	17:32	LB136371
	Nickel	3.06	+/-20	U	40.0	P	07/03/2025	17:32	LB136371
	Zinc	16.7	+/-20	U	40.0	P	07/03/2025	17:32	LB136371

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ENTACT

SDG No.: Q2463

Contract: ENTA05

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
-----------	---------	----------------	---------------------	--------------	------	---	------------------	------------------	---------------
