

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: ENTACT

SDG No.: Q2463

Contract: ENTA05

Lab Code: ACE

ICS Source: EPA

Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Cadmium	-1.03	1.0	103	-5	7	07/03/2025	12:30	LB136371
	Copper	17.1	2.0	855	-18	22	07/03/2025	12:30	LB136371
	Lead	-8.04			-12	12	07/03/2025	12:30	LB136371
	Nickel	11.4	2.0	570	-38	42	07/03/2025	12:30	LB136371
	Zinc	1.18			-40	40	07/03/2025	12:30	LB136371
ICSAB01	Cadmium	974	972	100	826	1120	07/03/2025	12:42	LB136371
	Copper	490	511	96	434	588	07/03/2025	12:42	LB136371
	Lead	40.3	49.0	82	37	61	07/03/2025	12:42	LB136371
	Nickel	994	954	104	810	1100	07/03/2025	12:42	LB136371
	Zinc	1040	952	109	809	1095	07/03/2025	12:42	LB136371