

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>ENTACT</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q2463</u>
Contract:	<u>ENTA05</u>			Lab Code:	<u>ACE</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q2463-01</u>	Client ID:	<u>TW-WTS-11DUP</u>
Percent Solids for Sample:	NA	Duplicate ID	Q2463-01DUP	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Cadmium	ug/L	20	3.00	U	3.00	U			P
Copper	ug/L	20	21.7		21.7		0		P
Lead	ug/L	20	5.99	J	6.69		11		P
Mercury	ug/L	20	0.19	J	0.18	J	3		CV
Nickel	ug/L	20	15.0	J	15.3	J	2		P
Zinc	ug/L	20	17.3	J	17.3	J	0		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>ENTACT</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q2463</u>
Contract:	<u>ENTA05</u>			Lab Code:	<u>ACE</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q2463-01MS</u>	Client ID:	<u>TW-WTS-11MSD</u>
Percent Solids for Sample:	NA	Duplicate ID	Q2463-01MSD	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Cadmium	ug/L	20	97.4		96.7		1		P
Copper	ug/L	20	171		168		2		P
Lead	ug/L	20	466		456		2		P
Mercury	ug/L	20	3.24		3.88		18		CV
Nickel	ug/L	20	264		259		2		P
Zinc	ug/L	20	124		122		2		P