

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: ENTACT

SDG No.: Q2463

Contract: ENTA05

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168677BS							
Cadmium	ug/L	100	95.4		95	80 - 120	P
Copper	ug/L	150	151		101	80 - 120	P
Lead	ug/L	500	469		94	80 - 120	P
Nickel	ug/L	250	245		98	80 - 120	P
Zinc	ug/L	100	97.5		98	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: ENTACT

SDG No.: Q2463

Contract: ENTA05

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168683BS Mercury	ug/L	4.0	3.97		99	80 - 120	CV