

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ENTA05

Lab Code: ACE

SDG NO.: Q2463

Continuing Calib Date: 07/02/2025

Initial Calibration Date(s): 06/11/2025

06/11/2025

Continuing Calib Time: 22:46

Initial Calibration Time(s): 10:40

19:07

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1254-1 (1)	5.57	5.57	5.47	5.67	0.00
Aroclor-1254-2 (2)	5.72	5.72	5.62	5.82	0.00
Aroclor-1254-3 (3)	6.12	6.13	6.03	6.23	0.01
Aroclor-1254-4 (4)	6.35	6.36	6.26	6.46	0.01
Aroclor-1254-5 (5)	6.77	6.78	6.68	6.88	0.01
Tetrachloro-m-xylene	3.68	3.68	3.58	3.78	0.01
Decachlorobiphenyl	8.71	8.71	8.61	8.81	0.00

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ENTA05

Lab Code: ACE

SDG NO.: Q2463

Continuing Calib Date: 07/02/2025

Initial Calibration Date(s): 06/11/2025

06/11/2025

Continuing Calib Time: 22:46

Initial Calibration Time(s): 10:40

19:07

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1254-1 (1)	5.55	5.55	5.45	5.65	0.00
Aroclor-1254-2 (2)	5.69	5.70	5.60	5.80	0.01
Aroclor-1254-3 (3)	6.10	6.10	6.00	6.20	0.01
Aroclor-1254-4 (4)	6.32	6.33	6.23	6.43	0.01
Aroclor-1254-5 (5)	6.74	6.74	6.64	6.84	0.00
Tetrachloro-m-xylene	3.67	3.67	3.57	3.77	0.00
Decachlorobiphenyl	8.66	8.66	8.56	8.76	0.00