

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENTA05
Lab Code: CHEM **SDG NO.:** Q2463
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 06/11/2025 06/11/2025

Client Sample No.: CCAL01 **Date Analyzed:** 07/02/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO111981.D **Time Analyzed:** 16:19

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.763	4.666	4.866	494.270	500.000	-1.1
Aroclor-1016-2	4.782	4.684	4.884	498.670	500.000	-0.3
Aroclor-1016-3	4.838	4.741	4.941	493.110	500.000	-1.4
Aroclor-1016-4	4.958	4.861	5.061	499.120	500.000	-0.2
Aroclor-1016-5	5.215	5.118	5.318	509.150	500.000	1.8
Aroclor-1260-1	6.253	6.157	6.357	497.320	500.000	-0.5
Aroclor-1260-2	6.442	6.346	6.546	530.770	500.000	6.2
Aroclor-1260-3	6.809	6.713	6.913	543.990	500.000	8.8
Aroclor-1260-4	7.069	6.972	7.172	529.570	500.000	5.9
Aroclor-1260-5	7.312	7.216	7.416	514.080	500.000	2.8
Decachlorobiphenyl	8.707	8.612	8.812	44.820	50.000	-10.4
Tetrachloro-m-xylene	3.676	3.577	3.777	53.460	50.000	6.9

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENTA05
Lab Code: CHEM **SDG NO.:** Q2463
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 06/11/2025 06/11/2025

Client Sample No.: CCAL01 **Date Analyzed:** 07/02/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO111981.D **Time Analyzed:** 16:19

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.746	4.651	4.851	531.680	500.000	6.3
Aroclor-1016-2	4.765	4.669	4.869	540.660	500.000	8.1
Aroclor-1016-3	4.940	4.844	5.044	533.100	500.000	6.6
Aroclor-1016-4	4.983	4.887	5.087	532.020	500.000	6.4
Aroclor-1016-5	5.195	5.099	5.299	523.140	500.000	4.6
Aroclor-1260-1	6.224	6.129	6.329	531.030	500.000	6.2
Aroclor-1260-2	6.412	6.316	6.516	552.990	500.000	10.6
Aroclor-1260-3	6.564	6.469	6.669	523.940	500.000	4.8
Aroclor-1260-4	7.035	6.939	7.139	515.020	500.000	3.0
Aroclor-1260-5	7.276	7.181	7.381	524.740	500.000	4.9
Decachlorobiphenyl	8.656	8.561	8.761	51.330	50.000	2.7
Tetrachloro-m-xylene	3.671	3.573	3.773	52.280	50.000	4.6