

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENTA05
Lab Code: CHEM **SDG NO.:** Q2463
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 06/11/2025 06/11/2025

Client Sample No.: CCAL03 **Date Analyzed:** 07/03/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112001.D **Time Analyzed:** 08:31

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.762	4.666	4.866	515.290	500.000	3.1
Aroclor-1016-2	4.780	4.684	4.884	529.370	500.000	5.9
Aroclor-1016-3	4.837	4.741	4.941	579.710	500.000	15.9
Aroclor-1016-4	4.957	4.861	5.061	518.540	500.000	3.7
Aroclor-1016-5	5.215	5.118	5.318	535.260	500.000	7.1
Aroclor-1260-1	6.253	6.157	6.357	500.990	500.000	0.2
Aroclor-1260-2	6.443	6.346	6.546	565.310	500.000	13.1
Aroclor-1260-3	6.809	6.713	6.913	548.790	500.000	9.8
Aroclor-1260-4	7.069	6.972	7.172	530.310	500.000	6.1
Aroclor-1260-5	7.312	7.216	7.416	526.610	500.000	5.3
Decachlorobiphenyl	8.707	8.612	8.812	44.760	50.000	-10.5
Tetrachloro-m-xylene	3.675	3.577	3.777	57.630	50.000	15.3

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENTA05
Lab Code: CHEM **SDG NO.:** Q2463
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 06/11/2025 06/11/2025

Client Sample No.: CCAL03 **Date Analyzed:** 07/03/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112001.D **Time Analyzed:** 08:31

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.746	4.651	4.851	544.630	500.000	8.9
Aroclor-1016-2	4.765	4.669	4.869	550.590	500.000	10.1
Aroclor-1016-3	4.940	4.844	5.044	539.290	500.000	7.9
Aroclor-1016-4	4.982	4.887	5.087	538.610	500.000	7.7
Aroclor-1016-5	5.194	5.099	5.299	527.640	500.000	5.5
Aroclor-1260-1	6.224	6.129	6.329	540.570	500.000	8.1
Aroclor-1260-2	6.412	6.316	6.516	585.650	500.000	17.1
Aroclor-1260-3	6.564	6.469	6.669	532.490	500.000	6.5
Aroclor-1260-4	7.034	6.939	7.139	511.290	500.000	2.3
Aroclor-1260-5	7.275	7.181	7.381	532.520	500.000	6.5
Decachlorobiphenyl	8.655	8.561	8.761	50.290	50.000	0.6
Tetrachloro-m-xylene	3.670	3.573	3.773	54.330	50.000	8.7