

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>ENTA05</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q2463</u>
Instrument ID:	<u>BNA_F</u>	Calibration Date/Time:	<u>07/03/2025 09:55</u>
Lab File ID:	<u>BF142987.D</u>	Init. Calib. Date(s):	<u>06/19/2025 06/19/2025</u>
EPA Sample No.:	<u>SSTDCCC040</u>	Init. Calib. Time(s):	<u>17:07 20:40</u>
GC Column:	<u>DB-UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
2-Fluorophenol	1.201	1.155		-3.8	
Phenol-d6	1.417	1.350		-4.7	
Phenol	1.575	1.488		-5.5	20.0
1,4-Dichlorobenzene	1.462	1.411		-3.5	20.0
Nitrobenzene-d5	0.365	0.383		4.9	
1,2,4-Trichlorobenzene	0.310	0.302		-2.6	
Naphthalene	0.979	0.960		-1.9	
2-Fluorobiphenyl	1.522	1.626		6.8	
2,4,6-Tribromophenol	0.206	0.200		-2.9	
Terphenyl-d14	1.447	1.388		-4.1	

All other compounds must meet a minimum RRF of 0.010.