

Surrogate Summary

SW-846

SDG No.: Q2463

Client: ENTACT

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB168716BL	PB168716BL	2-Fluorophenol	150	129	86		15 (23)	110 (138)
		Phenol-d6	150	129	86		15 (10)	110 (134)
		Nitrobenzene-d5	100	79.8	80		30 (67)	130 (132)
		2-Fluorobiphenyl	100	78.6	79		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	136	91		15 (44)	110 (137)
		Terphenyl-d14	100	88.2	88		30 (42)	130 (152)
PB168716BS	PB168716BS	2-Fluorophenol	150	128	85		15 (23)	110 (138)
		Phenol-d6	150	130	87		15 (10)	110 (134)
		Nitrobenzene-d5	100	82.0	82		30 (67)	130 (132)
		2-Fluorobiphenyl	100	82.2	82		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	138	92		15 (44)	110 (137)
		Terphenyl-d14	100	87.6	88		30 (42)	130 (152)
PB168716BSD	PB168716BSD	2-Fluorophenol	150	116	77		15 (23)	110 (138)
		Phenol-d6	150	116	78		15 (10)	110 (134)
		Nitrobenzene-d5	100	74.9	75		30 (67)	130 (132)
		2-Fluorobiphenyl	100	75.5	75		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	128	85		15 (44)	110 (137)
		Terphenyl-d14	100	81.6	82		30 (42)	130 (152)
Q2463-01	TW-WTS-11	2-Fluorophenol	150	56.6	38		15 (23)	110 (138)
		Phenol-d6	150	37.2	25		15 (10)	110 (134)
		Nitrobenzene-d5	100	70.8	71		30 (67)	130 (132)
		2-Fluorobiphenyl	100	69.8	70		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	107	71		15 (44)	110 (137)
		Terphenyl-d14	100	55.3	55		30 (42)	130 (152)