

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2463

Analytical Method: 8270E

Client: ENTACT

DataFile: BF143028.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Limits		
								Qual	Low	High	RPD
PB168716BSD	Phenol	50	40.3	ug/L	81	8			20 (66)	160 (118)	20 (20)
	1,4-Dichlorobenzene	50	40.3	ug/L	81	11			70 (76)	130 (103)	20 (20)
	1,2,4-Trichlorobenzene	50	42.2	ug/L	84	6			70 (41)	130 (115)	20 (20)
	Naphthalene	50	42.2	ug/L	84	8			70 (64)	130 (107)	20 (20)