

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: <u>Alliance</u>	Contract: <u>ENTA05</u>
Lab Code: <u>ACE</u>	SDG No.: <u>Q2463</u>
Instrument ID: <u>MSVOA_X</u>	Calibration Date(s): <u>06/17/2025</u> <u>06/17/2025</u>
Heated Purge: (Y/N) <u>N</u>	Calibration Time(s): <u>11:19</u> <u>17:18</u>
GC Column: <u>DB-624UI</u> ID: <u>0.18</u> (mm)	

LAB FILE ID:		RRF005 = VX046718.D		RRF020 = VX046719.D		RRF050 = VX046720.D		
		RRF100 = VX046721.D		RRF150 = VX046722.D		RRF001 = VX046725.D		
COMPOUND	RRF005	RRF020	RRF050	RRF100	RRF150	RRF001	RRF	% RSD
Methyl tert-butyl Ether	1.628	1.803	1.752	1.898	1.808	1.427	1.720	9.8
Carbon Tetrachloride	0.509	0.537	0.515	0.548	0.531	0.470	0.518	5.4
Chloroform	1.102	1.190	1.114	1.181	1.109	0.857	1.092	11.1
1,1,1-Trichloroethane	0.931	1.012	0.956	1.046	0.990	0.812	0.958	8.6
Benzene	1.431	1.477	1.417	1.509	1.427	1.218	1.413	7.2
Toluene	0.884	0.927	0.888	0.927	0.881	0.750	0.876	7.4
Tetrachloroethene	0.345	0.353	0.340	0.360	0.339	0.341	0.346	2.4
Ethyl Benzene	1.905	2.030	1.933	2.062	1.945	1.696	1.929	6.7
m/p-Xylenes	0.705	0.764	0.724	0.765	0.718	0.635	0.719	6.7
o-Xylene	0.686	0.729	0.692	0.739	0.698	0.498	0.673	13.1
1,2-Dichloroethane-d4	0.801	0.567	0.649	0.726	0.714		0.692	12.7
Dibromofluoromethane	0.362	0.267	0.320	0.354	0.351		0.331	11.9
Toluene-d8	1.342	0.973	1.144	1.250	1.234		1.189	11.7
4-Bromofluorobenzene	0.513	0.354	0.426	0.466	0.456		0.443	13.3

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.