

Surrogate Summary

SDG No.: Q2463

Client: ENTACT

Analytical Method: SW8260-Low

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q2463-01	TW-WTS-11	1,2-Dichloroethane-d4	50	48.1	96		70 (74)	130 (125)
		Dibromofluoromethane	50	45.5	91		70 (75)	130 (124)
		Toluene-d8	50	47.7	95		70 (86)	130 (113)
		4-Bromofluorobenzene	50	49.8	100		70 (77)	130 (121)
VX0701WBL01	VX0701WBL01	1,2-Dichloroethane-d4	50	49.7	99		70 (74)	130 (125)
		Dibromofluoromethane	50	51.1	102		70 (75)	130 (124)
		Toluene-d8	50	50.5	101		70 (86)	130 (113)
		4-Bromofluorobenzene	50	52.2	104		70 (77)	130 (121)
VX0701WBS01	VX0701WBS01	1,2-Dichloroethane-d4	50	42.7	85		70 (74)	130 (125)
		Dibromofluoromethane	50	46.8	94		70 (75)	130 (124)
		Toluene-d8	50	47.3	94		70 (86)	130 (113)
		4-Bromofluorobenzene	50	48.0	96		70 (77)	130 (121)
VX0701WBSD0	VX0701WBSD01	1,2-Dichloroethane-d4	50	45.2	90		70 (74)	130 (125)
		Dibromofluoromethane	50	48.3	97		70 (75)	130 (124)
		Toluene-d8	50	47.9	96		70 (86)	130 (113)
		4-Bromofluorobenzene	50	49.0	98		70 (77)	130 (121)

