

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q2463</u>	Analytical Method:	<u>SW8260-Low</u>
Client:	<u>ENTACT</u>	Datafile :	<u>VX046848.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0701WBSD01	Methyl tert-butyl Ether	20	18.3	ug/L	92	6		70 (78)	130 (114)	20 (20)
	Carbon Tetrachloride	20	18.6	ug/L	93	0		70 (77)	130 (113)	20 (15)
	Chloroform	20	19.7	ug/L	99	1		70 (79)	130 (113)	20 (20)
	1,1,1-Trichloroethane	20	18.4	ug/L	92	0		70 (80)	130 (108)	20 (20)
	Benzene	20	19.4	ug/L	97	0		70 (82)	130 (109)	20 (15)
	Toluene	20	19.9	ug/L	100	2		70 (82)	130 (110)	20 (16)
	Tetrachloroethene	20	18.9	ug/L	95	0		70 (67)	130 (123)	20 (15)
	Ethyl Benzene	20	18.9	ug/L	95	1		70 (83)	130 (109)	20 (16)
	m/p-Xylenes	40	39.2	ug/L	98	0		70 (82)	130 (110)	20 (15)
	o-Xylene	20	19.9	ug/L	100	1		70 (83)	130 (109)	20 (20)