

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: ENTA05
 Lab Code: ACE SDG NO.: Q2463
 Lab File ID: VX046844.D Date Analyzed: 07/01/2025
 Instrument ID: MSVOA_X Time Analyzed: 08:53
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	147192	5.56	233068	6.76	203219	10.05
UPPER LIMIT	294384	6.056	466136	7.263	406438	10.549
LOWER LIMIT	73596	5.056	116534	6.263	101610	9.549
EPA SAMPLE NO.						
TW-WTS-11	149182	5.56	256031	6.76	234680	10.06
VX0701WBL01	140918	5.56	239139	6.77	216898	10.06
VX0701WBS01	144669	5.56	236747	6.76	211980	10.05
VX0701WBSD01	134985	5.56	220693	6.76	201379	10.05

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.