

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q2467 SAS No.: Q2467 SDG NO.: Q2467

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 06/11/2025 06/11/2025

Client Sample No.: CCAL01 Date Analyzed: 07/02/2025

Lab Sample No.: AR1660CCC500 Data File : PO111967.D Time Analyzed: 10:56

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.763	4.666	4.866	499.470	500.000	-0.1
Aroclor-1016-2	4.781	4.684	4.884	523.980	500.000	4.8
Aroclor-1016-3	4.838	4.741	4.941	502.290	500.000	0.5
Aroclor-1016-4	4.958	4.861	5.061	505.900	500.000	1.2
Aroclor-1016-5	5.215	5.118	5.318	532.110	500.000	6.4
Aroclor-1260-1	6.253	6.157	6.357	524.450	500.000	4.9
Aroclor-1260-2	6.443	6.346	6.546	567.730	500.000	13.5
Aroclor-1260-3	6.809	6.713	6.913	592.080	500.000	18.4
Aroclor-1260-4	7.070	6.972	7.172	583.190	500.000	16.6
Aroclor-1260-5	7.312	7.216	7.416	566.850	500.000	13.4
Decachlorobiphenyl	8.708	8.612	8.812	47.330	50.000	-5.3
Tetrachloro-m-xylene	3.675	3.577	3.777	53.730	50.000	7.5

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q2467 SAS No.: Q2467 SDG NO.: Q2467

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/11/2025 06/11/2025

Client Sample No.: CCAL01 Date Analyzed: 07/02/2025

Lab Sample No.: AR1660CCC500 Data File : PO111967.D Time Analyzed: 10:56

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.746	4.651	4.851	524.840	500.000	5.0
Aroclor-1016-2	4.765	4.669	4.869	539.070	500.000	7.8
Aroclor-1016-3	4.940	4.844	5.044	527.680	500.000	5.5
Aroclor-1016-4	4.983	4.887	5.087	528.380	500.000	5.7
Aroclor-1016-5	5.194	5.099	5.299	521.700	500.000	4.3
Aroclor-1260-1	6.225	6.129	6.329	536.180	500.000	7.2
Aroclor-1260-2	6.413	6.316	6.516	547.830	500.000	9.6
Aroclor-1260-3	6.564	6.469	6.669	524.840	500.000	5.0
Aroclor-1260-4	7.034	6.939	7.139	510.430	500.000	2.1
Aroclor-1260-5	7.276	7.181	7.381	517.150	500.000	3.4
Decachlorobiphenyl	8.656	8.561	8.761	49.900	50.000	-0.2
Tetrachloro-m-xylene	3.671	3.573	3.773	52.020	50.000	4.0