



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

JPCL Engineering

Project Name: NOHO RFI No. SC64025 NYC

Project # N/A

Order ID # Q2473

Test Name: VOC-TCL

A. Number of Samples and Date of Receipt:

8 Solid samples were received on 07/01/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: Corrosivity, Cyanide, Herbicide, Ignitability, Mercury, Metals ICP-Group1, Metals ICP-TAL, METALS-TAL, PCB, PESTICIDE Group1, Pesticide-TCL, RCRA CHARACTERISTICS, Reactive Cyanide, Reactive Sulfide, SVOC-Chemtech Full, SVOC-TCL BNA -20, TCLP Extraction, TCLP Mercury, TCLP Metal, TCLP Metals Group3, TPH, TPH GC, VOC-TCL and VOC-TCLVOA-10. This data package contains results for VOC-TCL.

C. Analytical Techniques:

The analysis performed on instrument MSVOA_Y were done using GC column Rxi-624SIL MS 30m, 0.25mm, 1.4 um, Cat. #13868. The analysis of VOC-TCL was based on method 8260D.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria except for PIT#4 [Dibromofluoromethane - 61%], PIT#4RE [Dibromofluoromethane - 62%] sample was reanalyzed to confirm the failure and reported.

The Internal Standards Areas met the acceptable requirements except for PIT#4RE sample was reanalyzed to confirm the failure and reported.

The Retention Times were acceptable for all samples.

The RPD met criteria.

The Blank Spike met requirements for all samples.

The Blank Spike Duplicate met requirements for all samples.

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements.



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

The Continuous Calibration File ID VY022903.D met the requirements except for Acetone is failing high and associate sample having hit of acetone therefore sample was reanalyzed to confirm the failure and reported.

The Tuning criteria met requirements.

E. Additional Comments:

Samples for MS/MSD for VOC analysis were not provided with this set of samples. The Blank Spike Duplicate is reported with the data.

Trip Blank was not provided with this set of samples.

The soil samples results are based on a dry weight basis.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <20% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 20% for the Initial Calibration curve for SW-846 analysis.

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____