

Initial and Continuing Calibration Verification

Client: JPCL Engineering

SDG No.: Q2473

Project: NOHO RFI No. SC64025 NYC

RunNo.: LB136351

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: Corrosivity	ICV	pH	7.02	7	100	90-110	07/02/2025
Sample ID: Corrosivity	CCV1	pH	2.02	2.00	101	90-110	07/02/2025
Sample ID: Corrosivity	CCV2	pH	12.02	12.00	100	90-110	07/02/2025

Initial and Continuing Calibration Verification

Client: JPCL Engineering

SDG No.: Q2473

Project: NOHO RFI No. SC64025 NYC

RunNo.: LB136353

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Reactive Cyanide	mg/L	0.097	0.099	98	85-115	07/02/2025
Sample ID: CCV1 Reactive Cyanide	mg/L	0.24	0.25	96	90-110	07/02/2025
Sample ID: CCV2 Reactive Cyanide	mg/L	0.25	0.25	100	90-110	07/02/2025
Sample ID: CCV3 Reactive Cyanide	mg/L	0.25	0.25	100	90-110	07/02/2025

Initial and Continuing Calibration Verification

Client: JPCL Engineering

SDG No.: Q2473

Project: NOHO RFI No. SC64025 NYC

RunNo.: LB136358

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Cyanide	mg/L	0.096	0.099	97	90-110	07/02/2025
Sample ID: CCV1 Cyanide	mg/L	0.24	0.25	96	90-110	07/02/2025
Sample ID: CCV2 Cyanide	mg/L	0.24	0.25	96	90-110	07/02/2025
Sample ID: CCV3 Cyanide	mg/L	0.25	0.25	100	90-110	07/02/2025
Sample ID: CCV4 Cyanide	mg/L	0.25	0.25	100	90-110	07/02/2025