

Laboratory Control Sample Summary

Client: JPCL Engineering

SDG No.: Q2473

Project: NOHO RFI No. SC64025 NYC

Run No.: LB136358

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB168686BS							
Cyanide	mg/Kg	5	4.80		96	1	85-115	07/02/2025