

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: <u>JPCL Engineering</u>	SDG No.: <u>Q2473</u>
Contract: <u>JPCL01</u>	Lab Code: <u>ACE</u>
ICS Source: <u>EPA</u>	Instrument ID: <u>P4</u>

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Aluminum	241000	255000	94	216000	294000	07/07/2025	13:53	LB136390
	Antimony	5.00			-50	50	07/07/2025	13:53	LB136390
	Arsenic	-4.19			-20	20	07/07/2025	13:53	LB136390
	Barium	-3.69	6.0	62	-94	106	07/07/2025	13:53	LB136390
	Beryllium	0.58			-6	6	07/07/2025	13:53	LB136390
	Cadmium	-0.042	1.0	4	-5	7	07/07/2025	13:53	LB136390
	Calcium	231000	245000	94	208000	282000	07/07/2025	13:53	LB136390
	Chromium	53.6	52.0	103	42	62	07/07/2025	13:53	LB136390
	Cobalt	0.58			-30	30	07/07/2025	13:53	LB136390
	Copper	10.9	2.0	544	-18	22	07/07/2025	13:53	LB136390
	Iron	101000	101000	100	85600	116500	07/07/2025	13:53	LB136390
	Lead	-5.45			-12	12	07/07/2025	13:53	LB136390
	Magnesium	244000	255000	96	216000	294000	07/07/2025	13:53	LB136390
	Manganese	6.77	7.0	97	-13	27	07/07/2025	13:53	LB136390
	Nickel	15.9	2.0	795	-38	42	07/07/2025	13:53	LB136390
	Potassium	133			0	0	07/07/2025	13:53	LB136390
	Selenium	8.20			-20	20	07/07/2025	13:53	LB136390
	Silver	-6.96			-10	10	07/07/2025	13:53	LB136390
	Sodium	101			0	0	07/07/2025	13:53	LB136390
	Thallium	-0.81			-40	40	07/07/2025	13:53	LB136390
	Vanadium	4.26			-40	40	07/07/2025	13:53	LB136390
	Zinc	3.42			-40	40	07/07/2025	13:53	LB136390
ICSAB01	Aluminum	238000	247000	96	209000	285000	07/07/2025	14:10	LB136390
	Antimony	622	618	101	525	711	07/07/2025	14:10	LB136390
	Arsenic	99.1	104	95	88.4	120	07/07/2025	14:10	LB136390
	Barium	455	537	85	437	637	07/07/2025	14:10	LB136390
	Beryllium	472	495	95	420	570	07/07/2025	14:10	LB136390
	Cadmium	992	972	102	826	1120	07/07/2025	14:10	LB136390
	Calcium	227000	235000	97	199000	271000	07/07/2025	14:10	LB136390
	Chromium	558	542	103	460	624	07/07/2025	14:10	LB136390
	Cobalt	501	476	105	404	548	07/07/2025	14:10	LB136390
	Copper	495	511	97	434	588	07/07/2025	14:10	LB136390
	Iron	103000	99300	104	84400	114500	07/07/2025	14:10	LB136390
	Lead	41.5	49.0	85	37	61	07/07/2025	14:10	LB136390
	Magnesium	238000	248000	96	210000	286000	07/07/2025	14:10	LB136390
	Manganese	473	507	93	430	584	07/07/2025	14:10	LB136390
	Nickel	1010	954	106	810	1100	07/07/2025	14:10	LB136390
	Potassium	-260			0	0	07/07/2025	14:10	LB136390
	Selenium	53.0	46.0	115	26	66	07/07/2025	14:10	LB136390
	Silver	208	201	104	170	232	07/07/2025	14:10	LB136390
	Sodium	16.7			0	0	07/07/2025	14:10	LB136390
	Thallium	78.7	108	73	68	148	07/07/2025	14:10	LB136390

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: <u>JPCL Engineering</u>	SDG No.: <u>Q2473</u>
Contract: <u>JPCL01</u>	Lab Code: <u>ACE</u>
ICS Source: <u>EPA</u>	Instrument ID: <u>P4</u>

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSAB01	Vanadium	459	491	94	417	565	07/07/2025	14:10	LB136390
	Zinc	1050	952	110	809	1095	07/07/2025	14:10	LB136390
ICSA	Aluminum	240000	255000	94	216000	294000	07/07/2025	14:15	LB136390
	Antimony	9.95			-50	50	07/07/2025	14:15	LB136390
	Arsenic	-16.2			-20	20	07/07/2025	14:15	LB136390
	Barium	-48.6	6.0	810	-94	106	07/07/2025	14:15	LB136390
	Beryllium	-0.36			-6	6	07/07/2025	14:15	LB136390
	Cadmium	-2.96	1.0	296	-5	7	07/07/2025	14:15	LB136390
	Calcium	236000	245000	96	208000	282000	07/07/2025	14:15	LB136390
	Chromium	153	52.0	294	42	62	07/07/2025	14:15	LB136390
	Cobalt	1.25			-30	30	07/07/2025	14:15	LB136390
	Copper	24.1	2.0	1205	-18	22	07/07/2025	14:15	LB136390
	Iron	99500	101000	98	85600	116500	07/07/2025	14:15	LB136390
	Lead	0.64			-12	12	07/07/2025	14:15	LB136390
	Magnesium	244000	255000	96	216000	294000	07/07/2025	14:15	LB136390
	Manganese	11.2	7.0	160	-13	27	07/07/2025	14:15	LB136390
	Nickel	70.7	2.0	3535	-38	42	07/07/2025	14:15	LB136390
	Potassium	-7370			0	0	07/07/2025	14:15	LB136390
	Selenium	-53.4			-20	20	07/07/2025	14:15	LB136390
	Silver	-8.88			-10	10	07/07/2025	14:15	LB136390
	Sodium	-118			0	0	07/07/2025	14:15	LB136390
	Thallium	-40.5			-40	40	07/07/2025	14:15	LB136390
	Vanadium	8.77			-40	40	07/07/2025	14:15	LB136390
	Zinc	12.7			-40	40	07/07/2025	14:15	LB136390
ICSAB	Aluminum	244000	247000	99	209000	285000	07/07/2025	14:19	LB136390
	Antimony	588	618	95	525	711	07/07/2025	14:19	LB136390
	Arsenic	94.5	104	91	88.4	120	07/07/2025	14:19	LB136390
	Barium	474	537	88	437	637	07/07/2025	14:19	LB136390
	Beryllium	493	495	100	420	570	07/07/2025	14:19	LB136390
	Cadmium	970	972	100	826	1120	07/07/2025	14:19	LB136390
	Calcium	242000	235000	103	199000	271000	07/07/2025	14:19	LB136390
	Chromium	587	542	108	460	624	07/07/2025	14:19	LB136390
	Cobalt	489	476	103	404	548	07/07/2025	14:19	LB136390
	Copper	515	511	101	434	588	07/07/2025	14:19	LB136390
	Iron	102000	99300	103	84400	114500	07/07/2025	14:19	LB136390
	Lead	47.0	49.0	96	37	61	07/07/2025	14:19	LB136390
	Magnesium	250000	248000	101	210000	286000	07/07/2025	14:19	LB136390
	Manganese	490	507	97	430	584	07/07/2025	14:19	LB136390
	Nickel	1000	954	105	810	1100	07/07/2025	14:19	LB136390
	Potassium	-8010			0	0	07/07/2025	14:19	LB136390
	Selenium	19.1	46.0	42	26	66	07/07/2025	14:19	LB136390
	Silver	227	201	113	170	232	07/07/2025	14:19	LB136390

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: JPCL Engineering

SDG No.: Q2473

Contract: JPCL01

Lab Code: ACE

ICS Source: EPA

Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSAB	Sodium	-111			0	0	07/07/2025	14:19	LB136390
	Thallium	48.2	108	45	68	148	07/07/2025	14:19	LB136390
	Vanadium	527	491	107	417	565	07/07/2025	14:19	LB136390
	Zinc	1030	952	108	809	1095	07/07/2025	14:19	LB136390
ICSA01	Aluminum	245000	250000	98	216000	294000	07/09/2025	10:31	LB136407
	Antimony	23.3			-50	50	07/09/2025	10:31	LB136407
	Arsenic	5.46			-20	20	07/09/2025	10:31	LB136407
	Barium	8.02	6.0	134	-94	106	07/09/2025	10:31	LB136407
	Beryllium	1.50			-6	6	07/09/2025	10:31	LB136407
	Cadmium	-0.43	1.0	43	-5	7	07/09/2025	10:31	LB136407
	Calcium	242000	240000	101	208000	282000	07/09/2025	10:31	LB136407
	Chromium	52.2	52.0	100	42	62	07/09/2025	10:31	LB136407
	Cobalt	27.0			-30	30	07/09/2025	10:31	LB136407
	Copper	-4.74	2.0	237	-18	22	07/09/2025	10:31	LB136407
	Iron	96800	100000	97	85600	116500	07/09/2025	10:31	LB136407
	Lead	-5.64			-12	12	07/09/2025	10:31	LB136407
	Magnesium	242000	260000	93	216000	294000	07/09/2025	10:31	LB136407
	Manganese	3.37	7.0	48	-13	27	07/09/2025	10:31	LB136407
	Nickel	13.5	2.0	675	-38	42	07/09/2025	10:31	LB136407
	Potassium	10.6			0	0	07/09/2025	10:31	LB136407
	Selenium	-11.2			-20	20	07/09/2025	10:31	LB136407
	Silver	-2.41			-10	10	07/09/2025	10:31	LB136407
	Sodium	119			0	0	07/09/2025	10:31	LB136407
	Thallium	16.7			-40	40	07/09/2025	10:31	LB136407
	Vanadium	1.75			-40	40	07/09/2025	10:31	LB136407
	Zinc	10.9			-40	40	07/09/2025	10:31	LB136407
ICSAB01	Aluminum	245000	250000	98	209000	285000	07/09/2025	10:35	LB136407
	Antimony	621	620	100	525	711	07/09/2025	10:35	LB136407
	Arsenic	98.6	100	99	88.4	120	07/09/2025	10:35	LB136407
	Barium	529	540	98	437	637	07/09/2025	10:35	LB136407
	Beryllium	514	500	103	420	570	07/09/2025	10:35	LB136407
	Cadmium	984	970	101	826	1120	07/09/2025	10:35	LB136407
	Calcium	241000	230000	105	199000	271000	07/09/2025	10:35	LB136407
	Chromium	555	540	103	460	624	07/09/2025	10:35	LB136407
	Cobalt	506	480	105	404	548	07/09/2025	10:35	LB136407
	Copper	479	510	94	434	588	07/09/2025	10:35	LB136407
	Iron	98200	99000	99	84400	114500	07/09/2025	10:35	LB136407
	Lead	38.6	49.0	79	37	61	07/09/2025	10:35	LB136407
	Magnesium	243000	250000	97	210000	286000	07/09/2025	10:35	LB136407
	Manganese	498	510	98	430	584	07/09/2025	10:35	LB136407
	Nickel	966	950	102	810	1100	07/09/2025	10:35	LB136407
	Potassium	9.40			0	0	07/09/2025	10:35	LB136407

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: <u>JPCL Engineering</u>	SDG No.: <u>Q2473</u>
Contract: <u>JPCL01</u>	Lab Code: <u>ACE</u>
ICS Source: <u>EPA</u>	Instrument ID: <u>P5</u>

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSAB01	Selenium	40.2	46.0	87	26	66	07/09/2025	10:35	LB136407
	Silver	209	200	104	170	232	07/09/2025	10:35	LB136407
	Sodium	100			0	0	07/09/2025	10:35	LB136407
	Thallium	101	110	92	68	148	07/09/2025	10:35	LB136407
	Vanadium	489	490	100	417	565	07/09/2025	10:35	LB136407
	Zinc	1030	950	108	809	1095	07/09/2025	10:35	LB136407
ICSA	Aluminum	256000	250000	102	216000	294000	07/09/2025	10:39	LB136407
	Antimony	56.4			-50	50	07/09/2025	10:39	LB136407
	Arsenic	-38.5			-20	20	07/09/2025	10:39	LB136407
	Barium	3.35	6.0	56	-94	106	07/09/2025	10:39	LB136407
	Beryllium	3.59			-6	6	07/09/2025	10:39	LB136407
	Cadmium	-4.72	1.0	472	-5	7	07/09/2025	10:39	LB136407
	Calcium	253000	240000	105	208000	282000	07/09/2025	10:39	LB136407
	Chromium	30.5	52.0	59	42	62	07/09/2025	10:39	LB136407
	Cobalt	120			-30	30	07/09/2025	10:39	LB136407
	Copper	39.8	2.0	1990	-18	22	07/09/2025	10:39	LB136407
	Iron	108000	100000	108	85600	116500	07/09/2025	10:39	LB136407
	Lead	24.1			-12	12	07/09/2025	10:39	LB136407
	Magnesium	259000	260000	100	216000	294000	07/09/2025	10:39	LB136407
	Manganese	25.6	7.0	366	-13	27	07/09/2025	10:39	LB136407
	Nickel	-4.00	2.0	200	-38	42	07/09/2025	10:39	LB136407
	Potassium	59.1			0	0	07/09/2025	10:39	LB136407
	Selenium	91.3			-20	20	07/09/2025	10:39	LB136407
	Silver	-6.28			-10	10	07/09/2025	10:39	LB136407
	Sodium	-123			0	0	07/09/2025	10:39	LB136407
	Thallium	-21.0			-40	40	07/09/2025	10:39	LB136407
	Vanadium	-159			-40	40	07/09/2025	10:39	LB136407
	Zinc	21.5			-40	40	07/09/2025	10:39	LB136407
ICSAB	Aluminum	239000	250000	96	209000	285000	07/09/2025	10:44	LB136407
	Antimony	588	620	95	525	711	07/09/2025	10:44	LB136407
	Arsenic	95.9	100	96	88.4	120	07/09/2025	10:44	LB136407
	Barium	555	540	103	437	637	07/09/2025	10:44	LB136407
	Beryllium	497	500	99	420	570	07/09/2025	10:44	LB136407
	Cadmium	885	970	91	826	1120	07/09/2025	10:44	LB136407
	Calcium	236000	230000	103	199000	271000	07/09/2025	10:44	LB136407
	Chromium	527	540	98	460	624	07/09/2025	10:44	LB136407
	Cobalt	563	480	117	404	548	07/09/2025	10:44	LB136407
	Copper	517	510	101	434	588	07/09/2025	10:44	LB136407
	Iron	101000	99000	102	84400	114500	07/09/2025	10:44	LB136407
	Lead	15.6	49.0	32	37	61	07/09/2025	10:44	LB136407
	Magnesium	245000	250000	98	210000	286000	07/09/2025	10:44	LB136407
	Manganese	558	510	109	430	584	07/09/2025	10:44	LB136407

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: <u>JPCL Engineering</u>	SDG No.: <u>Q2473</u>
Contract: <u>JPCL01</u>	Lab Code: <u>ACE</u>
ICS Source: <u>EPA</u>	Instrument ID: <u>P5</u>

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSAB	Nickel	883	950	93	810	1100	07/09/2025	10:44	LB136407
	Potassium	-58.8			0	0	07/09/2025	10:44	LB136407
	Selenium	-15.1	46.0	33	26	66	07/09/2025	10:44	LB136407
	Silver	181	200	90	170	232	07/09/2025	10:44	LB136407
	Sodium	-8.12			0	0	07/09/2025	10:44	LB136407
	Thallium	109	110	99	68	148	07/09/2025	10:44	LB136407
	Vanadium	305	490	62	417	565	07/09/2025	10:44	LB136407
	Zinc	991	950	104	809	1095	07/09/2025	10:44	LB136407
ICSA01	Aluminum	247000	250000	99	216000	294000	07/15/2025	14:10	LB136486
	Antimony	21.6			-50	50	07/15/2025	14:10	LB136486
	Arsenic	-2.03			-20	20	07/15/2025	14:10	LB136486
	Barium	6.35	6.0	106	-94	106	07/15/2025	14:10	LB136486
	Beryllium	1.09			-6	6	07/15/2025	14:10	LB136486
	Cadmium	-1.17	1.0	117	-5	7	07/15/2025	14:10	LB136486
	Calcium	236000	240000	98	208000	282000	07/15/2025	14:10	LB136486
	Chromium	52.5	52.0	101	42	62	07/15/2025	14:10	LB136486
	Cobalt	3.22			-30	30	07/15/2025	14:10	LB136486
	Copper	-6.71	2.0	336	-18	22	07/15/2025	14:10	LB136486
	Iron	97400	100000	97	85600	116500	07/15/2025	14:10	LB136486
	Lead	-0.60			-12	12	07/15/2025	14:10	LB136486
	Magnesium	242000	260000	93	216000	294000	07/15/2025	14:10	LB136486
	Manganese	6.23	7.0	89	-13	27	07/15/2025	14:10	LB136486
	Nickel	3.30	2.0	165	-38	42	07/15/2025	14:10	LB136486
	Potassium	5.78			0	0	07/15/2025	14:10	LB136486
	Selenium	-7.22			-20	20	07/15/2025	14:10	LB136486
	Silver	-0.55			-10	10	07/15/2025	14:10	LB136486
	Sodium	113			0	0	07/15/2025	14:10	LB136486
	Thallium	22.2			-40	40	07/15/2025	14:10	LB136486
	Vanadium	3.62			-40	40	07/15/2025	14:10	LB136486
	Zinc	10.5			-40	40	07/15/2025	14:10	LB136486
ICSAB01	Aluminum	245000	250000	98	209000	285000	07/15/2025	14:21	LB136486
	Antimony	611	620	98	525	711	07/15/2025	14:21	LB136486
	Arsenic	92.3	100	92	88.4	120	07/15/2025	14:21	LB136486
	Barium	490	540	91	437	637	07/15/2025	14:21	LB136486
	Beryllium	488	500	98	420	570	07/15/2025	14:21	LB136486
	Cadmium	959	970	99	826	1120	07/15/2025	14:21	LB136486
	Calcium	234000	230000	102	199000	271000	07/15/2025	14:21	LB136486
	Chromium	529	540	98	460	624	07/15/2025	14:21	LB136486
	Cobalt	478	480	100	404	548	07/15/2025	14:21	LB136486
	Copper	462	510	91	434	588	07/15/2025	14:21	LB136486
	Iron	95400	99000	96	84400	114500	07/15/2025	14:21	LB136486
	Lead	48.6	49.0	99	37	61	07/15/2025	14:21	LB136486

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: <u>JPCL Engineering</u>	SDG No.: <u>Q2473</u>
Contract: <u>JPCL01</u>	Lab Code: <u>ACE</u>
ICS Source: <u>EPA</u>	Instrument ID: <u>P5</u>

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSAB01	Magnesium	240000	250000	96	210000	286000	07/15/2025	14:21	LB136486
	Manganese	477	510	94	430	584	07/15/2025	14:21	LB136486
	Nickel	952	950	100	810	1100	07/15/2025	14:21	LB136486
	Potassium	-2.22			0	0	07/15/2025	14:21	LB136486
	Selenium	38.5	46.0	84	26	66	07/15/2025	14:21	LB136486
	Silver	208	200	104	170	232	07/15/2025	14:21	LB136486
	Sodium	99.8			0	0	07/15/2025	14:21	LB136486
	Thallium	101	110	92	68	148	07/15/2025	14:21	LB136486
	Vanadium	467	490	95	417	565	07/15/2025	14:21	LB136486
	Zinc	977	950	103	809	1095	07/15/2025	14:21	LB136486
ICSA	Aluminum	246000	250000	98	216000	294000	07/15/2025	14:26	LB136486
	Antimony	38.5			-50	50	07/15/2025	14:26	LB136486
	Arsenic	-116			-20	20	07/15/2025	14:26	LB136486
	Barium	2.83	6.0	47	-94	106	07/15/2025	14:26	LB136486
	Beryllium	-0.20			-6	6	07/15/2025	14:26	LB136486
	Cadmium	-2.62	1.0	262	-5	7	07/15/2025	14:26	LB136486
	Calcium	238000	240000	99	208000	282000	07/15/2025	14:26	LB136486
	Chromium	46.1	52.0	89	42	62	07/15/2025	14:26	LB136486
	Cobalt	113			-30	30	07/15/2025	14:26	LB136486
	Copper	2.79	2.0	140	-18	22	07/15/2025	14:26	LB136486
	Iron	107000	100000	107	85600	116500	07/15/2025	14:26	LB136486
	Lead	20.6			-12	12	07/15/2025	14:26	LB136486
	Magnesium	247000	260000	95	216000	294000	07/15/2025	14:26	LB136486
	Manganese	59.8	7.0	854	-13	27	07/15/2025	14:26	LB136486
	Nickel	-0.74	2.0	37	-38	42	07/15/2025	14:26	LB136486
	Potassium	-695			0	0	07/15/2025	14:26	LB136486
	Selenium	21.5			-20	20	07/15/2025	14:26	LB136486
	Silver	-13.2			-10	10	07/15/2025	14:26	LB136486
	Sodium	-321			0	0	07/15/2025	14:26	LB136486
	Thallium	98.2			-40	40	07/15/2025	14:26	LB136486
	Vanadium	92.6			-40	40	07/15/2025	14:26	LB136486
	Zinc	8.05			-40	40	07/15/2025	14:26	LB136486
ICSAB	Aluminum	258000	250000	103	209000	285000	07/15/2025	14:30	LB136486
	Antimony	670	620	108	525	711	07/15/2025	14:30	LB136486
	Arsenic	-16.0	100	16	88.4	120	07/15/2025	14:30	LB136486
	Barium	558	540	103	437	637	07/15/2025	14:30	LB136486
	Beryllium	554	500	111	420	570	07/15/2025	14:30	LB136486
	Cadmium	989	970	102	826	1120	07/15/2025	14:30	LB136486
	Calcium	249000	230000	108	199000	271000	07/15/2025	14:30	LB136486
	Chromium	617	540	114	460	624	07/15/2025	14:30	LB136486
	Cobalt	627	480	131	404	548	07/15/2025	14:30	LB136486
	Copper	581	510	114	434	588	07/15/2025	14:30	LB136486

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: JPCL Engineering

SDG No.: Q2473

Contract: JPCL01

Lab Code: ACE

ICS Source: EPA

Instrument ID: P5

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSAB	Iron	112000	99000	113	84400	114500	07/15/2025	14:30	LB136486
	Lead	72.1	49.0	147	37	61	07/15/2025	14:30	LB136486
	Magnesium	259000	250000	104	210000	286000	07/15/2025	14:30	LB136486
	Manganese	622	510	122	430	584	07/15/2025	14:30	LB136486
	Nickel	1010	950	106	810	1100	07/15/2025	14:30	LB136486
	Potassium	-673			0	0	07/15/2025	14:30	LB136486
	Selenium	-40.1	46.0	87	26	66	07/15/2025	14:30	LB136486
	Silver	79.4	200	40	170	232	07/15/2025	14:30	LB136486
	Sodium	-616			0	0	07/15/2025	14:30	LB136486
	Thallium	175	110	159	68	148	07/15/2025	14:30	LB136486
	Vanadium	414	490	84	417	565	07/15/2025	14:30	LB136486
	Zinc	1070	950	113	809	1095	07/15/2025	14:30	LB136486