

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: JPCL Engineering

SDG No.: Q2473

Contract: JPCL01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168713BS							
Aluminum	mg/Kg	100	96.5		96	80 - 120	P
Antimony	mg/Kg	40.0	39.2		98	80 - 120	P
Arsenic	mg/Kg	40.0	38.3		96	80 - 120	P
Barium	mg/Kg	10.0	10.2		102	80 - 120	P
Beryllium	mg/Kg	10.0	9.58		96	80 - 120	P
Cadmium	mg/Kg	10.0	9.63		96	80 - 120	P
Calcium	mg/Kg	50.0	50.6	J	101	80 - 120	P
Chromium	mg/Kg	20.0	20.9		104	80 - 120	P
Cobalt	mg/Kg	10.0	9.78		98	80 - 120	P
Copper	mg/Kg	15.0	15.1		101	80 - 120	P
Iron	mg/Kg	150	151		101	80 - 120	P
Lead	mg/Kg	50.0	47.3		95	80 - 120	P
Magnesium	mg/Kg	100	96.0	J	96	80 - 120	P
Manganese	mg/Kg	10.0	9.86		99	80 - 120	P
Nickel	mg/Kg	25.0	25.2		101	80 - 120	P
Potassium	mg/Kg	500	484		97	80 - 120	P
Selenium	mg/Kg	100	94.9		95	80 - 120	P
Silver	mg/Kg	3.8	3.56		94	80 - 120	P
Sodium	mg/Kg	150	154		103	80 - 120	P
Thallium	mg/Kg	100	81.4		81	80 - 120	P
Vanadium	mg/Kg	15.0	14.2		95	80 - 120	P
Zinc	mg/Kg	10.0	10.3		103	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: JPCL Engineering

SDG No.: Q2473

Contract: JPCL01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168719BS Mercury	mg/Kg	0.25	0.27		107	80 - 120	CV