

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB136390

Review By	Janvi	Review On	7/8/2025 11:37:30 AM
Supervise By	jaswal	Supervise On	7/9/2025 2:58:20 AM
STD. NAME	STD REF.#		
ICAL Standard	MP86192,MP86193,MP86194,MP86195,MP86196,MP86212		
ICV Standard	MP86213		
CCV Standard	MP86216		
ICSA Standard	MP86214,MP86215		
CRI Standard	MP86212		
LCS Standard			
Chk Standard	MP86217,MP86218		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	07/07/25 12:25		Jaswal	OK
2	S1	S1	CAL2	07/07/25 12:30		Jaswal	OK
3	S2	S2	CAL3	07/07/25 12:34		Jaswal	OK
4	S3	S3	CAL4	07/07/25 12:38		Jaswal	OK
5	S4	S4	CAL5	07/07/25 12:42		Jaswal	OK
6	S5	S5	CAL6	07/07/25 12:47		Jaswal	OK
7	ICV01	ICV01	ICV	07/07/25 13:22		Jaswal	OK
8	LLICV01	LLICV01	LLICV	07/07/25 13:41		Jaswal	OK
9	ICB01	ICB01	ICB	07/07/25 13:45		Jaswal	OK
10	CRI01	CRI01	CRDL	07/07/25 13:49		Jaswal	OK
11	ICSA01	ICSA01	ICSA	07/07/25 13:53		Jaswal	OK
12	ICSAB01	ICSAB01	ICSAB	07/07/25 14:10		Jaswal	OK
13	ICSADL	ICSADL	ICSA	07/07/25 14:15		Jaswal	OK
14	ICSABDL	ICSABDL	ICSAB	07/07/25 14:19		Jaswal	OK
15	CCV01	CCV01	CCV	07/07/25 14:38		Jaswal	OK
16	CCB01	CCB01	CCB	07/07/25 14:42		Jaswal	OK
17	Q2491-01	EO-1-070225	SAM	07/07/25 14:46		Jaswal	OK
18	Q2472-01DL	PN#1DL	SAM	07/07/25 14:50	Straight 5x for all elements	Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB136390

Review By	Janvi	Review On	7/8/2025 11:37:30 AM
Supervise By	jaswal	Supervise On	7/9/2025 2:58:20 AM

STD. NAME	STD REF.#
ICAL Standard	MP86192,MP86193,MP86194,MP86195,MP86196,MP86212
ICV Standard	MP86213
CCV Standard	MP86216
ICSA Standard	MP86214,MP86215
CRI Standard	MP86212
LCS Standard	
Chk Standard	MP86217,MP86218

19	Q2478-01	WC-1	SAM	07/07/25 14:55		Jaswal	OK
20	Q2381-08	VNJ-207	SAM	07/07/25 14:59		Jaswal	OK
21	Q2398-09	M00-25-0179	SAM	07/07/25 15:03	Confirm TCLP (Pb)	Jaswal	Not Ok
22	Q2472-02	PN#1	SAM	07/07/25 15:07		Jaswal	OK
23	Q2474-01	FO-1	SAM	07/07/25 15:12		Jaswal	OK
24	Q2478-04	WC-1	SAM	07/07/25 15:16	Confirm TCLP (Ba)	Jaswal	Not Ok
25	Q2473-05	PIT#1	SAM	07/07/25 15:21		Jaswal	OK
26	Q2473-06	PIT#2	SAM	07/07/25 15:25		Jaswal	OK
27	CCV02	CCV02	CCV	07/07/25 15:30		Jaswal	OK
28	CCB02	CCB02	CCB	07/07/25 15:34		Jaswal	OK
29	Q2472-01DL2	PN#1DL2	SAM	07/07/25 15:38	NOT USE	Jaswal	Not Ok
30	Q2507-01	SU-04-07032025	SAM	07/07/25 15:42		Jaswal	OK
31	Q2493-04	WC-11	SAM	07/07/25 15:46		Jaswal	OK
32	Q2493-04DUP	WC-11DUP	DUP	07/07/25 15:51		Jaswal	OK
33	Q2493-04L	WC-11L	SD	07/07/25 15:55		Jaswal	OK
34	Q2493-04MS	WC-11MS	MS	07/07/25 15:59		Jaswal	OK
35	Q2493-04MSD	WC-11MSD	MSD	07/07/25 16:04		Jaswal	OK
36	Q2493-04A	WC-11A	PS	07/07/25 16:08		Jaswal	OK
37	Q2477-01	50728	SAM	07/07/25 16:16	NOT USE	Jaswal	Not Ok
38	CCV03	CCV03	CCV	07/07/25 16:33		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB136390

Review By	Janvi	Review On	7/8/2025 11:37:30 AM
Supervise By	jaswal	Supervise On	7/9/2025 2:58:20 AM
STD. NAME	STD REF.#		
ICAL Standard	MP86192,MP86193,MP86194,MP86195,MP86196,MP86212		
ICV Standard	MP86213		
CCV Standard	MP86216		
ICSA Standard	MP86214,MP86215		
CRI Standard	MP86212		
LCS Standard			
Chk Standard	MP86217,MP86218		

39	CCB03	CCB03	CCB	07/07/25 16:37		Jaswal	OK
40	PB168711BL	PB168711BL	MB	07/07/25 16:41		Jaswal	OK
41	PB168713BL	PB168713BL	MB	07/07/25 16:45		Jaswal	OK
42	PB168714BL	PB168714BL	MB	07/07/25 16:50		Jaswal	OK
43	PB168713BS	PB168713BS	LCS	07/07/25 17:01		Jaswal	OK
44	PB168714BS	PB168714BS	LCS	07/07/25 17:05		Jaswal	OK
45	Q2473-01	PIT#1	SAM	07/07/25 17:09		Jaswal	OK
46	PB168711BS	PB168711BS	LCS	07/07/25 17:14		Jaswal	OK
47	Q2473-03	PIT#3	SAM	07/07/25 17:18		Jaswal	OK
48	Q2473-04	PIT#4	SAM	07/07/25 17:22		Jaswal	OK
49	Q2484-01	TP-58	SAM	07/07/25 17:26		Jaswal	OK
50	CCV04	CCV04	CCV	07/07/25 17:31		Jaswal	OK
51	CCB04	CCB04	CCB	07/07/25 17:35		Jaswal	OK
52	Q2484-02	TP-57	SAM	07/07/25 17:39	CCV fail for TI	Jaswal	Not Ok
53	Q2484-03	TP-64	SAM	07/07/25 17:43	CCV fail for TI	Jaswal	Not Ok
54	Q2484-04	TP-107	SAM	07/07/25 17:47	CCV fail for TI	Jaswal	Not Ok
55	Q2484-05	TP-106	SAM	07/07/25 17:51	CCV fail for TI	Jaswal	Not Ok
56	Q2484-06	TP-104	SAM	07/08/25 01:19	CCV fail for TI	Jaswal	Not Ok
57	Q2484-06DUP	TP-104DUP	DUP	07/08/25 01:23	CCV fail for TI	Jaswal	Not Ok
58	Q2484-06L	TP-104L	SD	07/08/25 01:27	CCV fail for TI	Jaswal	Not Ok

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB136390

Review By	Janvi	Review On	7/8/2025 11:37:30 AM
Supervise By	jaswal	Supervise On	7/9/2025 2:58:20 AM
STD. NAME	STD REF.#		
ICAL Standard	MP86192,MP86193,MP86194,MP86195,MP86196,MP86212		
ICV Standard	MP86213		
CCV Standard	MP86216		
ICSA Standard	MP86214,MP86215		
CRI Standard	MP86212		
LCS Standard			
Chk Standard	MP86217,MP86218		

59	Q2484-06MS	TP-104MS	MS	07/08/25 01:31	CCV fail for TI	Jaswal	Not Ok
60	Q2484-06MSD	TP-104MSD	MSD	07/08/25 01:35	CCV fail for TI	Jaswal	Not Ok
61	Q2484-06A	TP-104A	PS	07/08/25 01:39	CCV fail for TI	Jaswal	Not Ok
62	CCV05	CCV05	CCV	07/08/25 01:43	CCV fail for TI	Jaswal	OK
63	CCB05	CCB05	CCB	07/08/25 01:47		Jaswal	OK
64	Q2487-09	G4(0-6)	SAM	07/08/25 01:51	CCV fail for K,TI	Jaswal	Not Ok
65	Q2487-10	G4(6-12)	SAM	07/08/25 01:56	CCV fail for K,TI	Jaswal	Not Ok
66	Q2487-11	G3(0-6)	SAM	07/08/25 02:00	CCV fail for K,TI	Jaswal	Not Ok
67	Q2487-12	G3(6-12)	SAM	07/08/25 02:05	CCV fail for K,TI	Jaswal	Not Ok
68	Q2487-13	G2(0-6)	SAM	07/08/25 02:09	CCV fail for K,TI	Jaswal	Not Ok
69	Q2487-14	G2(6-12)	SAM	07/08/25 02:13	CCV fail for K,TI	Jaswal	Not Ok
70	Q2487-15	G1(0-6)	SAM	07/08/25 02:18	CCV fail for K,TI	Jaswal	Not Ok
71	Q2487-16	G1(6-12)	SAM	07/08/25 02:22	CCV fail for K,TI	Jaswal	Not Ok
72	Q2493-01	WC-11	SAM	07/08/25 02:41	CCV fail for K,TI	Jaswal	Not Ok
73	Q2493-01DUP	WC-11DUP	DUP	07/08/25 02:45	CCV fail for K,TI	Jaswal	Not Ok
74	CCV06	CCV06	CCV	07/08/25 02:49	CCV fail for K,TI	Jaswal	OK
75	CCB06	CCB06	CCB	07/08/25 02:53		Jaswal	OK
76	Q2493-01L	WC-11L	SD	07/08/25 02:58	CCV fail for K,TI	Jaswal	Not Ok
77	Q2493-01MS	WC-11MS	MS	07/08/25 03:02	CCV fail for K,TI	Jaswal	Not Ok
78	Q2493-01MSD	WC-11MSD	MSD	07/08/25 03:06	CCV fail for K,TI	Jaswal	Not Ok

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB136390

Review By	Janvi	Review On	7/8/2025 11:37:30 AM
Supervise By	jaswal	Supervise On	7/9/2025 2:58:20 AM
STD. NAME	STD REF.#		
ICAL Standard	MP86192,MP86193,MP86194,MP86195,MP86196,MP86212		
ICV Standard	MP86213		
CCV Standard	MP86216		
ICSA Standard	MP86214,MP86215		
CRI Standard	MP86212		
LCS Standard			
Chk Standard	MP86217,MP86218		

79	Q2493-01A	WC-11A	PS	07/08/25 03:10	CCV fail for K,TI	Jaswal	Not Ok
80	Q2473-07	PIT#3	SAM	07/08/25 03:14	Not Use	Jaswal	Not Ok
81	Q2473-08	PIT#4	SAM	07/08/25 03:18	Not Use	Jaswal	Not Ok
82	PB168689TB	PB168689TB	MB	07/08/25 03:23	Not Use	Jaswal	Not Ok
83	Q2473-02	PIT#2	SAM	07/08/25 03:36	CCV fail for K,TI	Jaswal	Not Ok
84	CCV07	CCV07	CCV	07/08/25 03:40		Jaswal	OK
85	CCB07	CCB07	CCB	07/08/25 03:44		Jaswal	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB136391

Review By	MOHAN	Review On	7/9/2025 11:32:13 AM
Supervise By	jaswal	Supervise On	7/9/2025 11:32:22 AM
STD. NAME	STD REF.#		
ICAL Standard	MP86288,MP86289,MP86290,MP86291,MP86292,MP86293		
ICV Standard	MP86294		
CCV Standard	MP86296		
ICSA Standard			
CRI Standard	MP86298		
LCS Standard			
Chk Standard	MP86295,MP86297,MP8699,MP86304		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	07/08/25 08:35		MOHAN	OK
2	S0.2	S0.2	CAL2	07/08/25 08:37		MOHAN	OK
3	S2.5	S2.5	CAL3	07/08/25 08:39		MOHAN	OK
4	S5	S5	CAL4	07/08/25 08:42		MOHAN	OK
5	S7.5	S7.5	CAL5	07/08/25 08:44		MOHAN	OK
6	S10	S10	CAL6	07/08/25 08:52		MOHAN	OK
7	ICV19	ICV19	ICV	07/08/25 08:59		MOHAN	OK
8	ICB19	ICB19	ICB	07/08/25 09:04		MOHAN	OK
9	CCV61	CCV61	CCV	07/08/25 09:07		MOHAN	OK
10	CCB61	CCB61	CCB	07/08/25 09:12		MOHAN	OK
11	CRA	CRA	CRDL	07/08/25 09:14		MOHAN	OK
12	HighStd	HighStd	HIGH STD	07/08/25 09:19		MOHAN	OK
13	ChkStd	ChkStd	SAM	07/08/25 09:21		MOHAN	OK
14	PB168741BL	PB168741BL	MB	07/08/25 09:24		MOHAN	OK
15	PB168741BS	PB168741BS	LCS	07/08/25 09:30		MOHAN	OK
16	Q2473-05	PIT#1	SAM	07/08/25 09:33		MOHAN	OK
17	Q2473-06	PIT#2	SAM	07/08/25 09:35		MOHAN	OK
18	Q2473-07	PIT#3	SAM	07/08/25 09:37		MOHAN	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB136391

Review By	MOHAN	Review On	7/9/2025 11:32:13 AM
Supervise By	jaswal	Supervise On	7/9/2025 11:32:22 AM
STD. NAME	STD REF.#		
ICAL Standard	MP86288,MP86289,MP86290,MP86291,MP86292,MP86293		
ICV Standard	MP86294		
CCV Standard	MP86296		
ICSA Standard			
CRI Standard	MP86298		
LCS Standard			
Chk Standard	MP86295,MP86297,MP8699,MP86304		

19	Q2473-08	PIT#4	SAM	07/08/25 09:40		MOHAN	OK
20	Q2474-01	FO-1	SAM	07/08/25 09:42		MOHAN	OK
21	CCV62	CCV62	CCV	07/08/25 09:44		MOHAN	OK
22	CCB62	CCB62	CCB	07/08/25 09:50		MOHAN	OK
23	Q2478-04	WC-1	SAM	07/08/25 09:52		MOHAN	OK
24	Q2493-04	WC-11	SAM	07/08/25 09:55		MOHAN	OK
25	Q2493-04DUP	WC-11DUP	DUP	07/08/25 09:57		MOHAN	OK
26	Q2493-04MS	WC-11MS	MS	07/08/25 10:05		MOHAN	OK
27	Q2493-04MSD	WC-11MSD	MSD	07/08/25 10:07		MOHAN	OK
28	PB168742BL	PB168742BL	MB	07/08/25 10:12		MOHAN	OK
29	PB168742BS	PB168742BS	LCS	07/08/25 10:15		MOHAN	OK
30	Q2481-01	CC0627-AL	SAM	07/08/25 10:17		MOHAN	OK
31	Q2481-01DUP	CC0627-ALDUP	DUP	07/08/25 10:19		MOHAN	OK
32	Q2481-01MS	CC0627-ALMS	MS	07/08/25 10:21		MOHAN	OK
33	CCV63	CCV63	CCV	07/08/25 10:24		MOHAN	OK
34	CCB63	CCB63	CCB	07/08/25 10:26		MOHAN	OK
35	Q2481-01MSD	CC0627-ALMSD	MSD	07/08/25 10:28		MOHAN	OK
36	Q2481-02DL	CC0627-CLOXPLDL	SAM	07/08/25 10:32	Report Straight 10X	MOHAN	OK
37	Q2481-03DL	CC0625-OBXBLDL	SAM	07/08/25 10:34	Report Straight 10X	MOHAN	OK
38	Q2481-04DL	CC0627-AOXLDL	SAM	07/08/25 10:37	Report Straight 10X	MOHAN	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB136391

Review By	MOHAN	Review On	7/9/2025 11:32:13 AM
Supervise By	jaswal	Supervise On	7/9/2025 11:32:22 AM
STD. NAME	STD REF.#		
ICAL Standard	MP86288,MP86289,MP86290,MP86291,MP86292,MP86293		
ICV Standard	MP86294		
CCV Standard	MP86296		
ICSA Standard			
CRI Standard	MP86298		
LCS Standard			
Chk Standard	MP86295,MP86297,MP8699,MP86304		

39	Q2481-05	CC0625-NL	SAM	07/08/25 10:39		MOHAN	OK
40	Q2481-06DL	CC0267-OXPLDL	SAM	07/08/25 10:42	Report Straight 10X	MOHAN	OK
41	Q2481-07DL	CC0627-OXLDL	SAM	07/08/25 10:44	Report Straight 10X	MOHAN	OK
42	Q2481-08	CC0627-CLOXAL	SAM	07/08/25 10:46		MOHAN	OK
43	Q2481-09DL	CC0627-BLDL	SAM	07/08/25 10:49	Report Straight 10X	MOHAN	OK
44	Q2481-10	CC0627-SFBL	SAM	07/08/25 10:51		MOHAN	OK
45	CCV64	CCV64	CCV	07/08/25 10:53		MOHAN	OK
46	CCB64	CCB64	CCB	07/08/25 10:56		MOHAN	OK
47	PB168689TB	PB168689TB	MB	07/08/25 10:58		MOHAN	OK
48	PB168705TB	PB168705TB	MB	07/08/25 11:03		MOHAN	OK
49	Q2493-04L	WC-11L	SD	07/08/25 11:05		MOHAN	OK
50	Q2493-04A	WC-11A	PS	07/08/25 11:07		MOHAN	OK
51	Q2481-01L	CC0627-ALL	SD	07/08/25 11:10		MOHAN	OK
52	Q2481-01A	CC0627-ALA	PS	07/08/25 11:12		MOHAN	OK
53	CCV65	CCV65	CCV	07/08/25 11:14		MOHAN	OK
54	CCB65	CCB65	CCB	07/08/25 11:17		MOHAN	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB136407

Review By	Janvi	Review On	7/10/2025 7:19:58 PM
Supervise By	jaswal	Supervise On	7/10/2025 11:33:47 PM
STD. NAME	STD REF.#		
ICAL Standard	MP86192,MP86193,MP86194,MP86195,MP86196,MP86212		
ICV Standard	MP86219		
CCV Standard	MP86216		
ICSA Standard	MP86214,MP86220		
CRI Standard	MP86212		
LCS Standard			
Chk Standard	MP86217,MP86218		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	07/09/25 09:39		Janvi	OK
2	S1	S1	CAL2	07/09/25 09:44		Janvi	OK
3	S2	S2	CAL3	07/09/25 09:48		Janvi	OK
4	S3	S3	CAL4	07/09/25 09:52		Janvi	OK
5	S4	S4	CAL5	07/09/25 09:57		Janvi	OK
6	S5	S5	CAL6	07/09/25 10:01		Janvi	OK
7	ICV01	ICV01	ICV	07/09/25 10:05		Janvi	OK
8	LLICV01	LLICV01	LLICV	07/09/25 10:17		Janvi	OK
9	ICB01	ICB01	ICB	07/09/25 10:22		Janvi	OK
10	CRI01	CRI01	CRDL	07/09/25 10:26		Janvi	OK
11	ICSA01	ICSA01	ICSA	07/09/25 10:31		Janvi	OK
12	ICSAB01	ICSAB01	ICSAB	07/09/25 10:35		Janvi	OK
13	ICSADL	ICSADL	ICSA	07/09/25 10:39		Janvi	OK
14	ICSABDL	ICSABDL	ICSAB	07/09/25 10:44		Janvi	OK
15	CCV01	CCV01	CCV	07/09/25 10:48		Janvi	OK
16	CCB01	CCB01	CCB	07/09/25 10:57		Janvi	OK
17	Q2425-01	COMP-1	SAM	07/09/25 11:02		Janvi	OK
18	Q2425-02	COMP-2	SAM	07/09/25 11:06		Janvi	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB136407

Review By	Janvi	Review On	7/10/2025 7:19:58 PM
Supervise By	jaswal	Supervise On	7/10/2025 11:33:47 PM
STD. NAME	STD REF.#		
ICAL Standard	MP86192,MP86193,MP86194,MP86195,MP86196,MP86212		
ICV Standard	MP86219		
CCV Standard	MP86216		
ICSA Standard	MP86214,MP86220		
CRI Standard	MP86212		
LCS Standard			
Chk Standard	MP86217,MP86218		

19	Q2425-03	COMP-3	SAM	07/09/25 11:11		Janvi	OK
20	Q2425-03DUP	COMP-3DUP	DUP	07/09/25 11:15		Janvi	OK
21	Q2425-03L	COMP-3L	SD	07/09/25 11:20		Janvi	OK
22	Q2425-03MS	COMP-3MS	MS	07/09/25 11:24		Janvi	OK
23	Q2425-03MSD	COMP-3MSD	MSD	07/09/25 11:29		Janvi	OK
24	Q2425-03A	COMP-3A	PS	07/09/25 11:33	0.1ML OF M6004 AND M6013 IN 10ML SAMPLE.	Janvi	OK
25	PB168628BL	PB168628BL	MB	07/09/25 11:38		Janvi	OK
26	PB168628BS	PB168628BS	LCS	07/09/25 11:42		Janvi	OK
27	CCV02	CCV02	CCV	07/09/25 11:47		Janvi	OK
28	CCB02	CCB02	CCB	07/09/25 11:56		Janvi	OK
29	Q2477-01	50728	SAM	07/09/25 12:10	CCB Fail for AI	Janvi	Not Ok
30	Q2477-01DUP	50728DUP	DUP	07/09/25 12:14	CCB Fail for AI	Janvi	Not Ok
31	Q2477-01L	50728L	SD	07/09/25 12:18	CCB Fail for AI	Janvi	Not Ok
32	Q2477-01MS	50728MS	MS	07/09/25 12:23	CCB Fail for AI	Janvi	Not Ok
33	Q2477-01MSD	50728MSD	MSD	07/09/25 12:27	CCB Fail for AI	Janvi	Not Ok
34	Q2477-01A	50728A	PS	07/09/25 12:31	CCB Fail for AI	Janvi	Not Ok
35	Q2473-07	PIT#3	SAM	07/09/25 12:36	NOT USE	Janvi	Not Ok
36	Q2473-08	PIT#4	SAM	07/09/25 12:40	NOT USE	Janvi	Not Ok
37	PB168748BL	PB168748BL	MB	07/09/25 12:45	CCB Fail for AI	Janvi	Not Ok

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB136407

Review By	Janvi	Review On	7/10/2025 7:19:58 PM
Supervise By	jaswal	Supervise On	7/10/2025 11:33:47 PM

STD. NAME	STD REF.#
ICAL Standard	MP86192,MP86193,MP86194,MP86195,MP86196,MP86212
ICV Standard	MP86219
CCV Standard	MP86216
ICSA Standard	MP86214,MP86220
CRI Standard	MP86212
LCS Standard	
Chk Standard	MP86217,MP86218

38	PB168748BS	PB168748BS	LCS	07/09/25 12:49	CCB Fail for Al	Janvi	Not Ok
39	CCV03	CCV03	CCV	07/09/25 12:54		Janvi	OK
40	CCB03	CCB03	CCB	07/09/25 12:58	CCB Fail for Al	Janvi	OK
41	Q2481-01DL	CC0627-ALDL	SAM	07/09/25 13:06	CCB Fail for Ag, Straight 10x Dilution for all elements	Janvi	Not Ok
42	Q2481-01DUPDL	CC0627-ALDUPDL	DUP	07/09/25 13:11	CCB Fail for Ag, Straight 10x Dilution for all elements	Janvi	Not Ok
43	Q2481-011DL	CC0627-AL	SAM	07/09/25 13:16	CCB Fail for Ag, Straight 10x Dilution for all elements	Janvi	Not Ok
44	Q2481-01MSDL	CC0627-ALMSDL	MS	07/09/25 13:20	CCB Fail for Ag, Straight 10x Dilution for all elements	Janvi	Not Ok
45	Q2481-01MSDDL	CC0627-ALMSDDL	MSD	07/09/25 13:25	CCB Fail for Ag, Straight 10x Dilution for all elements	Janvi	Not Ok
46	Q2481-01ADL	CC0627-ALADL	PS	07/09/25 13:30	CCB Fail for Ag, Straight 10x Dilution for all elements	Janvi	Not Ok
47	Q2481-02DL	CC0627-CLOXPLDL	SAM	07/09/25 13:35	CCB Fail for Ag, Straight 10x Dilution for all elements	Janvi	Not Ok
48	Q2481-03DL	CC0625-0XBOLDL	SAM	07/09/25 13:40	CCB Fail for Ag, Straight 10x Dilution for all elements	Janvi	Not Ok
49	Q2481-03DL2	CC0625-0XBOLDL2	SAM	07/09/25 13:44	CCB Fail for Ag, Straight 10x Dilution for all elements	Janvi	Not Ok
50	Q2481-04DL	CC0627-AOXLDL	SAM	07/09/25 13:49	CCB Fail for Ag, Straight 10x Dilution for all elements	Janvi	Not Ok
51	CCV04	CCV04	CCV	07/09/25 14:05		Janvi	OK
52	CCB04	CCB04	CCB	07/09/25 14:28	CCB Fail for Ag	Janvi	OK
53	Q2481-06DL	CC0267-0XPLDL	SAM	07/09/25 14:33	CCB Fail for Ag, Straight 5x Dilution for all elements	Janvi	Not Ok

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB136407

Review By	Janvi	Review On	7/10/2025 7:19:58 PM
Supervise By	jaswal	Supervise On	7/10/2025 11:33:47 PM

STD. NAME	STD REF.#
ICAL Standard	MP86192,MP86193,MP86194,MP86195,MP86196,MP86212
ICV Standard	MP86219
CCV Standard	MP86216
ICSA Standard	MP86214,MP86220
CRI Standard	MP86212
LCS Standard	
Chk Standard	MP86217,MP86218

54	Q2481-07DL	CC0627-OXLDL	SAM	07/09/25 14:37	CCB Fail for Ag, Straight 5x Dilution for all elements	Janvi	Not Ok
55	Q2481-08	CC0627-CLOXAL	SAM	07/09/25 14:42	CCB Fail for Ag	Janvi	Not Ok
56	Q2481-09DL	CC0627-BLDL	SAM	07/09/25 14:46	CCB Fail for Ag, Straight 5x Dilution for all elements	Janvi	Not Ok
57	Q2481-10DL	CC0627-SFBLDL	SAM	07/09/25 14:51	CCB Fail for Ag, Straight 5x Dilution for all elements	Janvi	Not Ok
58	Q2481-05DL	CC0625-NLDL	SAM	07/09/25 14:56	CCB Fail for Ag, Straight 5x Dilution for all elements	Janvi	Not Ok
59	Q2509-01	AUD-25-0110-0111	SAM	07/09/25 15:00	CCB Fail for Ag	Janvi	Not Ok
60	Q2512-01	WATER TREATMENT	SAM	07/09/25 15:05	CCB Fail for Ag	Janvi	Not Ok
61	Q2488-01	RW8-SP100-2025070	SAM	07/09/25 15:10		Janvi	OK
62	PB168705TB	PB168705TB	MB	07/09/25 15:14	CCB Fail for Ag	Janvi	Not Ok
63	CCV05	CCV05	CCV	07/09/25 15:27		Janvi	OK
64	CCB05	CCB05	CCB	07/09/25 15:33	CCB Fail for Ag	Janvi	OK
65	Q2503-01	GCAP2W	SAM	07/09/25 15:37	CCB Fail for Ag	Janvi	Not Ok
66	Q2512-01DUP	WATER TREATMENT	DUP	07/09/25 15:43	CCB Fail for Ag	Janvi	Not Ok
67	Q2512-01L	WATER TREATMENT	SD	07/09/25 15:47	CCB Fail for Ag	Janvi	Not Ok
68	Q2512-01MS	WATER TREATMENT	MS	07/09/25 15:52	CCB Fail for Ag	Janvi	Not Ok
69	Q2512-01MSD	WATER TREATMENT	MSD	07/09/25 15:56	CCB Fail for Ag	Janvi	Not Ok
70	Q2512-01A	WATER TREATMENT	PS	07/09/25 16:01	CCB Fail for Ag	Janvi	Not Ok
71	PB168712BL	PB168712BL	MB	07/09/25 16:05	CCB Fail for Ag	Janvi	Not Ok
72	PB168712BS	PB168712BS	LCS	07/09/25 16:09	CCB Fail for Ag	Janvi	Not Ok

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB136407

Review By	Janvi	Review On	7/10/2025 7:19:58 PM
Supervise By	jaswal	Supervise On	7/10/2025 11:33:47 PM
STD. NAME	STD REF.#		
ICAL Standard	MP86192,MP86193,MP86194,MP86195,MP86196,MP86212		
ICV Standard	MP86219		
CCV Standard	MP86216		
ICSA Standard	MP86214,MP86220		
CRI Standard	MP86212		
LCS Standard			
Chk Standard	MP86217,MP86218		

73	CCV06	CCV06	CCV	07/09/25 16:39		Janvi	OK
74	CCB06	CCB06	CCB	07/09/25 16:43		Janvi	OK
75	Q2488-02	RW8-SP303-2025070	SAM	07/09/25 16:52		Janvi	OK
76	Q2487-15	G1(0-6)	SAM	07/09/25 17:14	Bad injection	Janvi	Not Ok
77	Q2484-02	TP-57	SAM	07/09/25 17:19		Janvi	OK
78	Q2484-03	TP-64	SAM	07/09/25 17:23		Janvi	OK
79	Q2484-04	TP-107	SAM	07/09/25 17:28		Janvi	OK
80	Q2484-05	TP-106	SAM	07/09/25 17:32		Janvi	OK
81	Q2484-06	TP-104	SAM	07/09/25 17:37	NOT USE	Janvi	Not Ok
82	Q2484-06DUP	TP-104DUP	DUP	07/09/25 17:41	NOT USE	Janvi	Not Ok
83	Q2484-06L	TP-104L	SD	07/09/25 17:46	NOT USE	Janvi	Not Ok
84	CCV07	CCV07	CCV	07/09/25 17:50		Janvi	OK
85	CCB07	CCB07	CCB	07/09/25 17:54		Janvi	OK
86	Q2484-06MS	TP-104MS	MS	07/09/25 17:59	Bad injection	Janvi	Not Ok
87	Q2484-06MSD	TP-104MSD	MSD	07/09/25 18:03	Bad injection	Janvi	Not Ok
88	Q2484-06A	TP-104A	PS	07/09/25 18:08	NOT USE	Janvi	Not Ok
89	Q2487-09	G4(0-6)	SAM	07/09/25 18:12		Janvi	OK
90	Q2487-10	G4(6-12)	SAM	07/09/25 18:17		Janvi	OK
91	Q2487-11	G3(0-6)	SAM	07/09/25 18:21	Zn high	Janvi	Dilution
92	Q2487-12	G3(6-12)	SAM	07/09/25 18:26		Janvi	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB136407

Review By	Janvi	Review On	7/10/2025 7:19:58 PM
Supervise By	jaswal	Supervise On	7/10/2025 11:33:47 PM
STD. NAME	STD REF.#		
ICAL Standard	MP86192,MP86193,MP86194,MP86195,MP86196,MP86212		
ICV Standard	MP86219		
CCV Standard	MP86216		
ICSA Standard	MP86214,MP86220		
CRI Standard	MP86212		
LCS Standard			
Chk Standard	MP86217,MP86218		

93	Q2487-13	G2(0-6)	SAM	07/09/25 18:31		Janvi	OK
94	PB168689TB	PB168689TB	MB	07/09/25 18:35		Janvi	OK
95	Q2473-02	PIT#2	SAM	07/09/25 18:40		Janvi	OK
96	CCV08	CCV08	CCV	07/09/25 18:54		Janvi	OK
97	CCB08	CCB08	CCB	07/09/25 18:58		Janvi	OK
98	Q2487-16	G1(6-12)	SAM	07/09/25 19:03	Zn high	Janvi	Dilution
99	Q2503-01DL	GCAP2WDL	SAM	07/09/25 19:08	NOT USE	Janvi	Not Ok
100	PB168740BL	PB168740BL	MB	07/09/25 19:12	LCS fail for K	Janvi	Not Ok
101	PB168740BS	PB168740BS	LCS	07/09/25 19:16	LCS fail for K	Janvi	Not Ok
102	Q2503-05	GCAP2A	SAM	07/09/25 19:21		Janvi	OK
103	Q2513-01	HR-2-070325	SAM	07/09/25 19:25		Janvi	OK
104	Q2513-03	HR-3-070325	SAM	07/09/25 19:30		Janvi	OK
105	Q2514-01	TP-92	SAM	07/09/25 19:34		Janvi	OK
106	Q2514-02	TP-93	SAM	07/09/25 19:39		Janvi	OK
107	Q2514-03	TP-94	SAM	07/09/25 19:43		Janvi	OK
108	CCV09	CCV09	CCV	07/09/25 19:48		Janvi	OK
109	CCB09	CCB09	CCB	07/09/25 19:52		Janvi	OK
110	Q2514-04	TP-96	SAM	07/09/25 19:56		Janvi	OK
111	Q2514-05	TP-97	SAM	07/09/25 20:01		Janvi	OK
112	Q2514-06	TP-103	SAM	07/09/25 20:05		Janvi	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB136407

Review By	Janvi	Review On	7/10/2025 7:19:58 PM
Supervise By	jaswal	Supervise On	7/10/2025 11:33:47 PM
STD. NAME	STD REF.#		
ICAL Standard	MP86192,MP86193,MP86194,MP86195,MP86196,MP86212		
ICV Standard	MP86219		
CCV Standard	MP86216		
ICSA Standard	MP86214,MP86220		
CRI Standard	MP86212		
LCS Standard			
Chk Standard	MP86217,MP86218		

113	Q2514-07	TP-36	SAM	07/09/25 20:10		Janvi	OK
114	Q2514-08	TP-78	SAM	07/09/25 20:14		Janvi	OK
115	Q2514-09	TP-81	SAM	07/09/25 20:19		Janvi	OK
116	Q2514-10	TP-90	SAM	07/09/25 20:23		Janvi	OK
117	Q2515-01	WC-1	SAM	07/09/25 20:28	NOT USE	Janvi	Not Ok
118	Q2515-01DUP	WC-1DUP	DUP	07/09/25 20:32	NOT USE	Janvi	Not Ok
119	Q2515-01L	WC-1L	SD	07/09/25 20:37	NOT USE	Janvi	Not Ok
120	CCV10	CCV10	CCV	07/09/25 20:41		Janvi	OK
121	CCB10	CCB10	CCB	07/09/25 20:45		Janvi	OK
122	Q2515-01MS	WC-1MS	MS	07/09/25 20:50	Bad Injection	Janvi	Not Ok
123	Q2515-01MSD	WC-1MSD	MSD	07/09/25 20:54	NOT USE	Janvi	Not Ok
124	Q2515-01A	WC-1A	PS	07/09/25 20:59	NOT USE	Janvi	Not Ok
125	Q2487-14	G2(6-12)	SAM	07/09/25 21:03	Zn high	Janvi	Dilution
126	Q2487-15DL	G1(0-6)DL	SAM	07/09/25 21:08	NOT USE	Janvi	Not Ok
127	CCV11	CCV11	CCV	07/09/25 21:12		Janvi	OK
128	CCB11	CCB11	CCB	07/09/25 21:16		Janvi	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB136434

Review By	jaswal	Review On	7/12/2025 3:14:54 AM
Supervise By		Supervise On	
STD. NAME	STD REF.#		
ICAL Standard	MP86192,MP86193,MP86194,MP86195,MP86196,MP86212		
ICV Standard	MP86219		
CCV Standard	MP86216		
ICSA Standard	MP86214,MP86220		
CRI Standard	MP86212		
LCS Standard			
Chk Standard	MP86217,MP86218		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	07/10/25 16:46		Janvi	OK
2	S1	S1	CAL2	07/10/25 16:50		Janvi	OK
3	S2	S2	CAL3	07/10/25 16:55		Janvi	OK
4	S3	S3	CAL4	07/10/25 16:59		Janvi	OK
5	S4	S4	CAL5	07/10/25 17:03		Janvi	OK
6	S5	S5	CAL6	07/10/25 17:07		Janvi	OK
7	ICV01	ICV01	ICV	07/10/25 18:39	ICV fail for K,Ag (200.7) (95-105)	Janvi	OK
8	LLICV01	LLICV01	LLICV	07/10/25 18:51		Janvi	OK
9	ICB01	ICB01	ICB	07/10/25 18:55		Janvi	OK
10	CRI01	CRI01	CRDL	07/10/25 19:01		Janvi	OK
11	ICSA01	ICSA01	ICSA	07/10/25 19:11		Janvi	OK
12	ICSAB01	ICSAB01	ICSAB	07/10/25 19:15		Janvi	OK
13	ICSADL	ICSADL	ICSA	07/10/25 19:20		Janvi	OK
14	ICSABDL	ICSABDL	ICSAB	07/10/25 19:24		Janvi	OK
15	CCV01	CCV01	CCV	07/10/25 19:43		Janvi	OK
16	CCB01	CCB01	CCB	07/10/25 19:54		Janvi	OK
17	Q2532-01	001-WILLETS-PT-BL	SAM	07/10/25 19:58		Janvi	OK
18	Q2532-02	002-35th-Ave(May)	SAM	07/10/25 20:03		Janvi	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB136434

Review By	jaswal	Review On	7/12/2025 3:14:54 AM
Supervise By		Supervise On	
STD. NAME	STD REF.#		
ICAL Standard	MP86192,MP86193,MP86194,MP86195,MP86196,MP86212		
ICV Standard	MP86219		
CCV Standard	MP86216		
ICSA Standard	MP86214,MP86220		
CRI Standard	MP86212		
LCS Standard			
Chk Standard	MP86217,MP86218		

19	Q2533-01	001 WILLETS PT BLV	SAM	07/10/25 20:08		Janvi	OK
20	Q2533-02	002-35th-Ave(JUNE)	SAM	07/10/25 20:12		Janvi	OK
21	Q2509-01	AUD-25-0110-0111	SAM	07/10/25 20:17		Janvi	OK
22	Q2512-01	WATER TREATMENT	SAM	07/10/25 20:21		Janvi	OK
23	Q2520-24	A-24	SAM	07/10/25 20:26		Janvi	OK
24	Q2520-24DUP	A-24DUP	DUP	07/10/25 20:30		Janvi	OK
25	Q2520-24L	A-24L	SD	07/10/25 20:35		Janvi	OK
26	Q2520-24MS	A-24MS	MS	07/10/25 20:55		Janvi	OK
27	CCV02	CCV02	CCV	07/10/25 20:59		Janvi	OK
28	CCB02	CCB02	CCB	07/10/25 21:04		Janvi	OK
29	Q2520-24MSD	A-24MSD	MSD	07/10/25 21:09		Janvi	OK
30	Q2520-24A	A-24A	PS	07/10/25 21:14		Janvi	OK
31	PB168772BL	PB168772BL	MB	07/10/25 21:18	LCS fail for Fe	Janvi	Not Ok
32	Q2512-01DUP	WATER TREATMENT	DUP	07/10/25 21:27	K oversaturated	Janvi	Dilution
33	PB168772BS	PB168772BS	LCS	07/10/25 21:31	LCS fail for Fe	Janvi	Not Ok
34	Q2512-01L	WATER TREATMENT	SD	07/10/25 21:36		Janvi	OK
35	Q2512-01MS	WATER TREATMENT	MS	07/10/25 21:50	K oversaturated	Janvi	Dilution
36	Q2512-01MSD	WATER TREATMENT	MSD	07/10/25 21:59	K oversaturated	Janvi	Dilution
37	Q2512-01A	WATER TREATMENT	PS	07/10/25 22:03	K oversaturated	Janvi	Dilution
38	PB168705TB	PB168705TB	MB	07/10/25 22:07		Janvi	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB136434

Review By	jaswal	Review On	7/12/2025 3:14:54 AM
Supervise By		Supervise On	
STD. NAME	STD REF.#		
ICAL Standard	MP86192,MP86193,MP86194,MP86195,MP86196,MP86212		
ICV Standard	MP86219		
CCV Standard	MP86216		
ICSA Standard	MP86214,MP86220		
CRI Standard	MP86212		
LCS Standard			
Chk Standard	MP86217,MP86218		

39	CCV03	CCV03	CCV	07/10/25 22:32		Janvi	OK
40	CCB03	CCB03	CCB	07/10/25 22:36		Janvi	OK
41	Q2477-01	50728	SAM	07/10/25 22:41	confirm wt/vol	Janvi	OK
42	Q2477-01DUP	50728DUP	DUP	07/10/25 22:45		Janvi	OK
43	Q2477-01L	50728L	SD	07/10/25 22:49		Janvi	OK
44	Q2477-01MS	50728MS	MS	07/10/25 22:54		Janvi	OK
45	Q2477-01MSD	50728MSD	MSD	07/10/25 22:58		Janvi	OK
46	Q2477-01A	50728A	PS	07/10/25 23:02		Janvi	OK
47	Q2473-07	PIT#3	SAM	07/10/25 23:07		Janvi	OK
48	Q2473-08	PIT#4	SAM	07/10/25 23:11		Janvi	OK
49	PB168748BL	PB168748BL	MB	07/10/25 23:16		Janvi	OK
50	PB168748BS	PB168748BS	LCS	07/10/25 23:20		Janvi	OK
51	CCV04	CCV04	CCV	07/10/25 23:29		Janvi	OK
52	CCB04	CCB04	CCB	07/10/25 23:53		Janvi	OK
53	Q2512-01DL	WATER TREATMENT	SAM	07/10/25 23:57	5x for K	Janvi	Confirms
54	Q2512-01DUPDL	WATER TREATMENT	DUP	07/11/25 00:02	5x for K	Janvi	Confirms
55	Q2512-01LDL	WATER TREATMENT	SD	07/11/25 00:07		Janvi	OK
56	Q2512-01MSDL	WATER TREATMENT	MS	07/11/25 00:11	5x for K	Janvi	Confirms
57	Q2512-01MSDDL	WATER TREATMENT	MSD	07/11/25 00:16	5x for K	Janvi	Confirms
58	Q2512-01ADL	WATER TREATMENT	PS	07/11/25 00:20	5x for K	Janvi	Confirms

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB136434

Review By	jaswal	Review On	7/12/2025 3:14:54 AM
Supervise By		Supervise On	
STD. NAME	STD REF.#		
ICAL Standard	MP86192,MP86193,MP86194,MP86195,MP86196,MP86212		
ICV Standard	MP86219		
CCV Standard	MP86216		
ICSA Standard	MP86214,MP86220		
CRI Standard	MP86212		
LCS Standard			
Chk Standard	MP86217,MP86218		

59	PB168712BL	PB168712BL	MB	07/11/25 00:25		Janvi	OK
60	PB168712BS	PB168712BS	LCS	07/11/25 00:30		Janvi	OK
61	PB168740BL	PB168740BL	MB	07/11/25 00:34		Janvi	OK
62	PB168740BS	PB168740BS	LCS	07/11/25 00:38		Janvi	OK
63	CCV05	CCV05	CCV	07/11/25 00:47		Janvi	OK
64	CCB05	CCB05	CCB	07/11/25 00:51		Janvi	OK
65	PB168739BL	PB168739BL	MB	07/11/25 00:56		Janvi	OK
66	PB168739BS	PB168739BS	LCS	07/11/25 01:00		Janvi	OK
67	Q2481-01DL	CC0627-ALDL	SAM	07/11/25 01:04	Straight 10x for All elements	Janvi	OK
68	Q2481-01DUPDL	CC0627-ALDUPDL	DUP	07/11/25 01:09	Straight 10x for All elements	Janvi	OK
69	Q2481-01LDL	CC0627-ALLDL	SD	07/11/25 01:14	Straight 50x for All elements	Janvi	OK
70	Q2481-01MSDL	CC0627-ALMSDL	MS	07/11/25 01:19	Straight 10x for All elements	Janvi	OK
71	Q2481-01MSDDL	CC0627-ALMSDDL	MSD	07/11/25 01:24	Straight 10x for All elements	Janvi	OK
72	Q2481-01ADL	CC0627-ALADL	PS	07/11/25 01:29	Straight 10x for All elements	Janvi	OK
73	Q2481-02DL	CC0627-CLOXPLDL	SAM	07/11/25 01:34	Straight 10x for All elements	Janvi	OK
74	Q2481-03DL	CC0625-0XBLDL	SAM	07/11/25 01:39	Straight 10x for All elements	Janvi	OK
75	CCV06	CCV06	CCV	07/11/25 01:47		Janvi	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB136434

Review By	jaswal	Review On	7/12/2025 3:14:54 AM
Supervise By		Supervise On	
STD. NAME	STD REF.#		
ICAL Standard	MP86192,MP86193,MP86194,MP86195,MP86196,MP86212		
ICV Standard	MP86219		
CCV Standard	MP86216		
ICSA Standard	MP86214,MP86220		
CRI Standard	MP86212		
LCS Standard			
Chk Standard	MP86217,MP86218		

76	CCB06	CCB06	CCB	07/11/25 02:23		Janvi	OK
77	Q2481-04DL	CC0627-AOXLDL	SAM	07/11/25 02:27	Straight 5x for All elements	Janvi	OK
78	Q2481-05DL	CC0625-NLDL	SAM	07/11/25 02:32	Straight 5x for All elements	Janvi	OK
79	Q2481-06DL	CC0267-OXPLDL	SAM	07/11/25 02:36	Straight 5x for All elements	Janvi	OK
80	Q2481-07DL	CC0627-OXLDL	SAM	07/11/25 02:41	Straight 5x for All elements	Janvi	OK
81	Q2481-08	CC0627-CLOXAL	SAM	07/11/25 02:45		Janvi	OK
82	Q2481-09DL	CC0627-BLDL	SAM	07/11/25 02:50	Straight 10x for All elements	Janvi	OK
83	Q2481-10DL	CC0627-SFBLDL	SAM	07/11/25 02:55	Straight 5x for All elements	Janvi	OK
84	Q2481-03DL2	CC0625-OXBOLDL2	SAM	07/11/25 02:59	NOT USE	Janvi	Not Ok
85	Q2487-11DL	G3(0-6)DL	SAM	07/11/25 03:04	5x for Zn, Still Zn high	Janvi	Dilution
86	Q2487-16DL	G1(6-12)DL	SAM	07/11/25 03:08	5x for Zn, Still Zn high	Janvi	Dilution
87	CCV07	CCV07	CCV	07/11/25 03:17		Janvi	OK
88	CCB07	CCB07	CCB	07/11/25 03:26		Janvi	OK
89	Q2487-14DL	G2(6-12)DL	SAM	07/11/25 03:30	5x for Zn, Still Zn high	Janvi	Confirms
90	Q2520-21	A-21	SAM	07/11/25 03:34		Janvi	OK
91	Q2520-22	A-22	SAM	07/11/25 03:39		Janvi	OK
92	Q2520-23	A-23	SAM	07/11/25 03:43		Janvi	OK
93	PB168771BL	PB168771BL	MB	07/11/25 03:48		Janvi	OK
94	PB168771BS	PB168771BS	LCS	07/11/25 03:52		Janvi	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB136434

Review By	jaswal	Review On	7/12/2025 3:14:54 AM
Supervise By		Supervise On	
STD. NAME	STD REF.#		
ICAL Standard	MP86192,MP86193,MP86194,MP86195,MP86196,MP86212		
ICV Standard	MP86219		
CCV Standard	MP86216		
ICSA Standard	MP86214,MP86220		
CRI Standard	MP86212		
LCS Standard			
Chk Standard	MP86217,MP86218		

95	Q2520-01	A-1	SAM	07/11/25 03:56		Janvi	OK
96	Q2520-02	A-2	SAM	07/11/25 04:01		Janvi	OK
97	Q2520-03	A-3	SAM	07/11/25 04:05		Janvi	OK
98	Q2520-24RE	A-24RE	SAM	07/11/25 04:10	Not Use	Janvi	Not Ok
99	CCV08	CCV08	CCV	07/11/25 04:18		Janvi	OK
100	CCB08	CCB08	CCB	07/11/25 04:27	CCB fail for Ag	Janvi	OK
101	Q2520-05	A-5	SAM	07/11/25 04:31		Janvi	OK
102	Q2520-06	A-6	SAM	07/11/25 04:36	Bad injection	Janvi	Not Ok
103	Q2520-07	A-7	SAM	07/11/25 04:40		Janvi	OK
104	Q2520-08	A-8	SAM	07/11/25 04:44		Janvi	OK
105	Q2520-09	A-9	SAM	07/11/25 04:49		Janvi	OK
106	Q2520-10	A-10	SAM	07/11/25 04:53		Janvi	OK
107	Q2520-11	A-11	SAM	07/11/25 04:58		Janvi	OK
108	Q2520-12	A-12	SAM	07/11/25 05:02		Janvi	OK
109	Q2520-13	A-13	SAM	07/11/25 05:06		Janvi	OK
110	Q2520-14	A-14	SAM	07/11/25 05:11		Janvi	OK
111	CCV09	CCV09	CCV	07/11/25 05:19		Janvi	OK
112	CCB09	CCB09	CCB	07/11/25 05:24		Janvi	OK
113	Q2520-15	A-15	SAM	07/11/25 05:33		Janvi	OK
114	Q2520-16	A-16	SAM	07/11/25 05:37		Janvi	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB136434

Review By	jaswal	Review On	7/12/2025 3:14:54 AM
Supervise By		Supervise On	
STD. NAME	STD REF.#		
ICAL Standard	MP86192,MP86193,MP86194,MP86195,MP86196,MP86212		
ICV Standard	MP86219		
CCV Standard	MP86216		
ICSA Standard	MP86214,MP86220		
CRI Standard	MP86212		
LCS Standard			
Chk Standard	MP86217,MP86218		

115	Q2520-17	A-17	SAM	07/11/25 05:41		Janvi	OK
116	Q2520-18	A-18	SAM	07/11/25 05:46		Janvi	OK
117	Q2520-19	A-19	SAM	07/11/25 05:50		Janvi	OK
118	Q2520-20	A-20	SAM	07/11/25 05:55		Janvi	OK
119	Q2520-20DUP	A-20DUP	DUP	07/11/25 05:59		Janvi	OK
120	Q2520-20L	A-20L	SD	07/11/25 06:03		Janvi	OK
121	Q2520-20MS	A-20MS	MS	07/11/25 06:08		Janvi	OK
122	CCV10	CCV10	CCV	07/11/25 06:16		Janvi	OK
123	CCB10	CCB10	CCB	07/11/25 06:21		Janvi	OK
124	Q2520-20MSD	A-20MSD	MSD	07/11/25 06:30		Janvi	OK
125	Q2520-20A	A-20A	PS	07/11/25 06:34		Janvi	OK
126	CCV11	CCV11	CCV	07/11/25 06:43		Janvi	OK
127	CCB11	CCB11	CCB	07/11/25 06:47		Janvi	OK