

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: JPCL Engineering

SDG No.: Q2473

Contract: JPCL01

Lab Code: ACE

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Arsenic	23.0	20.0	115	65 - 135	P	07/07/2025	13:49	LB136390
	Barium	97.5	100	98	65 - 135	P	07/07/2025	13:49	LB136390
	Cadmium	7.08	6.0	118	65 - 135	P	07/07/2025	13:49	LB136390
	Chromium	11.8	10.0	118	65 - 135	P	07/07/2025	13:49	LB136390
	Lead	13.7	12.0	114	65 - 135	P	07/07/2025	13:49	LB136390
	Selenium	19.2	20.0	96	65 - 135	P	07/07/2025	13:49	LB136390
	Silver	10.6	10.0	106	65 - 135	P	07/07/2025	13:49	LB136390
CRA	Mercury	0.25	0.2	125	70 - 130	CV	07/08/2025	09:14	LB136391
CRI01	Arsenic	14.2	20.0	71	65 - 135	P	07/09/2025	10:26	LB136407
	Barium	104	100	104	65 - 135	P	07/09/2025	10:26	LB136407
	Cadmium	5.51	6.0	92	65 - 135	P	07/09/2025	10:26	LB136407
	Chromium	10.7	10.0	107	65 - 135	P	07/09/2025	10:26	LB136407
	Lead	11.5	12.0	96	65 - 135	P	07/09/2025	10:26	LB136407
	Selenium	22.8	20.0	114	65 - 135	P	07/09/2025	10:26	LB136407
	Silver	11.6	10.0	116	65 - 135	P	07/09/2025	10:26	LB136407
CRI01	Arsenic	20.6	20.0	103	65 - 135	P	07/10/2025	19:01	LB136434
	Barium	96.3	100	96	65 - 135	P	07/10/2025	19:01	LB136434
	Cadmium	5.33	6.0	89	65 - 135	P	07/10/2025	19:01	LB136434
	Chromium	10.5	10.0	105	65 - 135	P	07/10/2025	19:01	LB136434
	Lead	12.4	12.0	103	65 - 135	P	07/10/2025	19:01	LB136434
	Selenium	16.7	20.0	84	65 - 135	P	07/10/2025	19:01	LB136434
	Silver	10.5	10.0	105	65 - 135	P	07/10/2025	19:01	LB136434