

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: JPCL Engineering

SDG No.: Q2473

Instrument: CV1

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB168689TB		WATER		Batch Number:	PB168741		Prep Date:	07/07/2025	
	Mercury	0.76	<2	U	2.00	CV	07/08/2025	10:58	LB136391
Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB168741BL		WATER		Batch Number:	PB168741		Prep Date:	07/07/2025	
	Mercury	0.093	<0.2	J	0.20	CV	07/08/2025	09:24	LB136391

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: JPCL Engineering

SDG No.: Q2473

Instrument: P4

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB168711BL	WATER			Batch Number:	PB168711		Prep Date:	07/03/2025	
	Arsenic	25.6	<50	U	100	P	07/07/2025	16:41	LB136390
	Barium	72.8	<250	U	500	P	07/07/2025	16:41	LB136390
	Cadmium	2.50	<15	U	30.0	P	07/07/2025	16:41	LB136390
	Chromium	10.6	<25	U	50.0	P	07/07/2025	16:41	LB136390
	Lead	11.5	<30	U	60.0	P	07/07/2025	16:41	LB136390
	Selenium	48.2	<50	U	100	P	07/07/2025	16:41	LB136390
	Silver	8.10	<25	U	50.0	P	07/07/2025	16:41	LB136390

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: JPCL Engineering

SDG No.: Q2473

Instrument: P5

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB168689TB	WATER			Batch Number:	PB168711		Prep Date:	07/03/2025	
	Arsenic	25.6	<50	U	100	P	07/09/2025	18:35	LB136407
	Barium	72.8	<250	U	500	P	07/09/2025	18:35	LB136407
	Cadmium	2.50	<15	U	30.0	P	07/09/2025	18:35	LB136407
	Chromium	10.6	<25	U	50.0	P	07/09/2025	18:35	LB136407
	Lead	18.7	<30	J	60.0	P	07/09/2025	18:35	LB136407
	Selenium	48.2	<50	U	100	P	07/09/2025	18:35	LB136407
	Silver	8.10	<25	U	50.0	P	07/09/2025	18:35	LB136407