

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: JPCL Engineering

SDG No.: Q2473

Contract: JPCL01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168711BS							
Arsenic	ug/L	4000	3910		98	80 - 120	P
Barium	ug/L	1000	1010		101	80 - 120	P
Cadmium	ug/L	1000	976		98	80 - 120	P
Chromium	ug/L	2000	2120		106	80 - 120	P
Lead	ug/L	5000	4790		96	80 - 120	P
Selenium	ug/L	10000	9700		97	80 - 120	P
Silver	ug/L	380	364		96	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: JPCL Engineering

SDG No.: Q2473

Contract: JPCL01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168741BS Mercury	ug/L	4.0	3.55		89	80 - 120	CV