

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** JPCL01
Lab Code: ACE **SDG NO.:** Q2473
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 06/18/2025 06/18/2025

Client Sample No.: CCAL01 **Date Analyzed:** 07/07/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS030917.D **Time Analyzed:** 10:01

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.652	9.568	9.768	748.000	712.500	5.0
2,4,5-TP(Silvex)	9.358	9.275	9.475	696.660	712.500	-2.2
2,4-D	8.472	8.386	8.586	674.060	705.000	-4.4
2,4-DB	10.229	10.147	10.347	737.940	712.500	3.6
2,4-DCAA	7.337	7.249	7.449	690.210	750.000	-8.0
DICAMBA	7.526	7.439	7.639	672.540	705.000	-4.6
DICHLORPROP	8.240	8.153	8.353	630.010	705.000	-10.6
Dinoseb	11.446	11.366	11.566	693.780	705.000	-1.6

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** JPCL01
Lab Code: ACE **SDG NO.:** Q2473
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 06/18/2025 06/18/2025

Client Sample No.: CCAL01 **Date Analyzed:** 07/07/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS030917.D **Time Analyzed:** 10:01

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	10.354	10.259	10.459	711.460	712.500	-0.1
2,4,5-TP(Silvex)	9.929	9.833	10.033	701.470	712.500	-1.5
2,4-D	9.024	8.927	9.127	668.430	705.000	-5.2
2,4-DB	10.922	10.827	11.027	689.590	712.500	-3.2
2,4-DCAA	7.767	7.670	7.870	717.630	750.000	-4.3
DICAMBA	7.969	7.872	8.072	695.230	705.000	-1.4
DICHLORPROP	8.688	8.591	8.791	656.420	705.000	-6.9
Dinoseb	11.307	11.211	11.411	670.630	705.000	-4.9