

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** JPCL01
Lab Code: ACE **SDG NO.:** Q2473
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 06/18/2025 06/18/2025

Client Sample No.: CCAL05 **Date Analyzed:** 07/08/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS030944.D **Time Analyzed:** 09:27

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.651	9.568	9.768	770.730	712.500	8.2
2,4,5-TP(Silvex)	9.357	9.275	9.475	720.850	712.500	1.2
2,4-D	8.470	8.386	8.586	697.260	705.000	-1.1
2,4-DB	10.227	10.147	10.347	791.680	712.500	11.1
2,4-DCAA	7.336	7.249	7.449	733.720	750.000	-2.2
DICAMBA	7.525	7.439	7.639	690.450	705.000	-2.1
DICHLORPROP	8.238	8.153	8.353	650.710	705.000	-7.7
Dinoseb	11.444	11.366	11.566	691.050	705.000	-2.0

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** JPCL01
Lab Code: ACE **SDG NO.:** Q2473
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 06/18/2025 06/18/2025

Client Sample No.: CCAL05 **Date Analyzed:** 07/08/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS030944.D **Time Analyzed:** 09:27

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	10.353	10.259	10.459	690.590	712.500	-3.1
2,4,5-TP(Silvex)	9.927	9.833	10.033	686.060	712.500	-3.7
2,4-D	9.022	8.927	9.127	643.990	705.000	-8.7
2,4-DB	10.921	10.827	11.027	675.660	712.500	-5.2
2,4-DCAA	7.765	7.670	7.870	691.710	750.000	-7.8
DICAMBA	7.967	7.872	8.072	676.030	705.000	-4.1
DICHLORPROP	8.686	8.591	8.791	652.190	705.000	-7.5
Dinoseb	11.305	11.211	11.411	642.510	705.000	-8.9