

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** JPCL01
Lab Code: ACE **SDG NO.:** Q2473
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 06/18/2025 06/18/2025

Client Sample No.: CCAL06 **Date Analyzed:** 07/08/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS030956.D **Time Analyzed:** 16:05

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.654	9.568	9.768	763.100	712.500	7.1
2,4,5-TP(Silvex)	9.359	9.275	9.475	711.640	712.500	-0.1
2,4-D	8.473	8.386	8.586	693.790	705.000	-1.6
2,4-DB	10.231	10.147	10.347	786.470	712.500	10.4
2,4-DCAA	7.338	7.249	7.449	797.680	750.000	6.4
DICAMBA	7.527	7.439	7.639	688.250	705.000	-2.4
DICHLORPROP	8.241	8.153	8.353	643.890	705.000	-8.7
Dinoseb	11.449	11.366	11.566	696.870	705.000	-1.2

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** JPCL01
Lab Code: ACE **SDG NO.:** Q2473
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 06/18/2025 06/18/2025

Client Sample No.: CCAL06 **Date Analyzed:** 07/08/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS030956.D **Time Analyzed:** 16:05

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	10.361	10.259	10.459	683.490	712.500	-4.1
2,4,5-TP(Silvex)	9.935	9.833	10.033	676.910	712.500	-5.0
2,4-D	9.028	8.927	9.127	670.510	705.000	-4.9
2,4-DB	10.929	10.827	11.027	668.780	712.500	-6.1
2,4-DCAA	7.770	7.670	7.870	683.440	750.000	-8.9
DICAMBA	7.973	7.872	8.072	663.860	705.000	-5.8
DICHLORPROP	8.693	8.591	8.791	640.710	705.000	-9.1
Dinoseb	11.315	11.211	11.411	644.320	705.000	-8.6