

CALIBRATION VERIFICATION SUMMARY

Contract: JPCL01

Lab Code: CHEM Case No.: Q2473 SAS No.: Q2473 SDG NO.: Q2473

Continuing Calib Date: 07/02/2025 Initial Calibration Date(s): 07/01/2025 07/01/2025

Continuing Calib Time: 19:36 Initial Calibration Time(s): 14:04 21:30

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1254-1	(1)	6.49	6.49	6.39	6.59	0.00
Aroclor-1254-2	(2)	6.70	6.71	6.61	6.81	0.01
Aroclor-1254-3	(3)	7.07	7.07	6.97	7.17	0.00
Aroclor-1254-4	(4)	7.35	7.35	7.25	7.45	0.00
Aroclor-1254-5	(5)	7.76	7.77	7.67	7.87	0.01
Tetrachloro-m-xylene		4.49	4.49	4.39	4.59	0.00
Decachlorobiphenyl		10.18	10.18	10.08	10.28	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: JPCL01

Lab Code: CHEM Case No.: Q2473 SAS No.: Q2473 SDG NO.: Q2473

Continuing Calib Date: 07/02/2025 Initial Calibration Date(s): 07/01/2025 07/01/2025

Continuing Calib Time: 19:36 Initial Calibration Time(s): 14:04 21:30

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1254-1	(1)	5.66	5.66	5.56	5.76	0.00
Aroclor-1254-2	(2)	5.81	5.81	5.71	5.91	0.00
Aroclor-1254-3	(3)	6.21	6.21	6.11	6.31	0.00
Aroclor-1254-4	(4)	6.44	6.44	6.34	6.54	0.00
Aroclor-1254-5	(5)	6.86	6.86	6.76	6.96	0.00
Tetrachloro-m-xylene		3.78	3.78	3.68	3.88	0.00
Decachlorobiphenyl		8.79	8.79	8.69	8.89	0.00