

CALIBRATION VERIFICATION SUMMARY

Contract: JPCL01

Lab Code: CHEM Case No.: Q2473 SAS No.: Q2473 SDG NO.: Q2473

Continuing Calib Date: 07/03/2025 Initial Calibration Date(s): 07/01/2025 07/01/2025

Continuing Calib Time: 00:31 Initial Calibration Time(s): 14:04 21:30

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1248-1 (1)	5.64	5.64	5.54	5.74	0.00
Aroclor-1248-2 (2)	5.91	5.91	5.81	6.01	0.00
Aroclor-1248-3 (3)	6.11	6.11	6.01	6.21	0.00
Aroclor-1248-4 (4)	6.51	6.51	6.41	6.61	0.00
Aroclor-1248-5 (5)	6.55	6.55	6.45	6.65	0.00
Tetrachloro-m-xylene	4.49	4.49	4.39	4.59	0.00
Decachlorobiphenyl	10.18	10.18	10.08	10.28	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: JPCL01

Lab Code: CHEM Case No.: Q2473 SAS No.: Q2473 SDG NO.: Q2473

Continuing Calib Date: 07/03/2025 Initial Calibration Date(s): 07/01/2025 07/01/2025

Continuing Calib Time: 00:31 Initial Calibration Time(s): 14:04 21:30

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1248-1	(1)	4.86	4.86	4.76	4.96	0.00
Aroclor-1248-2	(2)	5.10	5.10	5.00	5.20	0.00
Aroclor-1248-3	(3)	5.14	5.14	5.04	5.24	0.00
Aroclor-1248-4	(4)	5.31	5.31	5.21	5.41	0.00
Aroclor-1248-5	(5)	5.70	5.70	5.60	5.80	0.00
Tetrachloro-m-xylene		3.78	3.79	3.69	3.89	0.01
Decachlorobiphenyl		8.79	8.79	8.69	8.89	0.01